

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102222\
 Data File : PQ059778.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2022 00:28
 Operator : YP\AJ
 Sample : N4955-08
 Misc : AR1268 LOQ 50 PPB
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOQ-WATER-02-QT4-2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 13:21:41 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ102122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 21 08:02:10 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.483	2.829	249.7E6	179.2E6	22.084	22.483
2) SA Decachlor...	8.650	7.600	145.2E6	219.0E6	26.566	26.063
Target Compounds						
41) L9 AR-1268-1	7.543	6.562	43195377	60683425	45.636	45.533
42) L9 AR-1268-2	7.622	6.624	38424727	50802847	45.015	41.629
43) L9 AR-1268-3	7.803	6.827	32054887	43851634	46.345	41.735
44) L9 AR-1268-4	8.122	7.114	13054418	17950428	45.335	40.940
45) L9 AR-1268-5	8.416	7.382	91400630	138.4E6	44.842	40.888

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102222\
Data File : PQ059778.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2022 00:28
Operator : YP\AJ
Sample : N4955-08
Misc : AR1268 LOQ 50 PPB
ALS Vial : 55 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT4-2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 13:21:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ102122.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 21 08:02:10 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

