

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102318\
 Data File : PQ033472.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Oct 2018 15:06
 Operator : SM\SJ
 Sample : J5164-07
 Misc : LOD AR1254 25PPB
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOD-MDL-WATER-01-QT4-2018

Manual Integrations
 APPROVED

Sohil
 10/26/2018 11:06:59 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 24 03:14:56 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 20 01:08:39 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.564	3.845	52288602	37104159	20.257	20.967
2) SA Decachlor...	10.391	8.892	36642129	30282964	17.860	18.299
Target Compounds						
26) L6 AR-1254-1	6.594	5.739	2791015	2760813	23.632m	21.557
27) L6 AR-1254-2	6.815	5.885	3871537	2665649	20.771m	23.918
28) L6 AR-1254-3	7.183	6.291	4388884	3591247	21.269	18.484
29) L6 AR-1254-4	7.469	6.516	3511305	2426663	23.002	19.755m
30) L6 AR-1254-5	7.889	6.935	2850723	3484962	16.915	20.520

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102318\
Data File : PQ033472.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Oct 2018 15:06
Operator : SM\SJ
Sample : J5164-07
Misc : LOD AR1254 25PPB
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2018

Manual Integrations
APPROVED

Sohil
10/26/2018 11:06:59 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 03:14:56 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101918.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 20 01:08:39 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

