

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102325\
 Data File : PQ071116.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Oct 2025 19:41
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 24 02:42:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 11 01:55:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.399	2.776	244.5E6	267.4E6	56.202	52.643
2)	SA Decachlor...	8.497	7.532	176.7E6	295.7E6	60.861	58.551
Target Compounds							
26)	L6 AR-1254-1	5.293	4.531	100.0E6	160.1E6	509.503	474.795
27)	L6 AR-1254-2	5.507	4.677	154.7E6	142.1E6	532.629	478.125
28)	L6 AR-1254-3	5.855	5.058	163.3E6	231.0E6	531.440	494.539
29)	L6 AR-1254-4	6.135	5.284	110.8E6	143.7E6	504.860	486.686
30)	L6 AR-1254-5	6.544	5.688	128.6E6	214.1E6	554.673	509.321

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102325\
Data File : PQ071116.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Oct 2025 19:41
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 02:42:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

