

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : DTX102920.M
 Title : GC Extractables
 Last Update : Fri Oct 30 04:27:50 2020
 Response Via : Initial Calibration

Calibration Files

500 =PD059982.D 1000 =PD059980.D 750 =PD059981.D
 250 =PD059983.D 100 =PD059984.D

	Compound	500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	9.140	9.149	8.945	9.014	9.339	9.117	E5	1.65
2)	Toxaphene-1	1.210	1.143	1.120	1.248	1.480	1.240	E4	11.56
3)	Toxaphene-2	1.959	1.933	1.872	1.950	2.069	1.957	E4	3.66
4)	Toxaphene-3	2.058	2.019	2.022	2.050	2.241	2.078	E4	4.47
5)	Toxaphene-4	7.206	7.153	6.876	7.033	7.210	7.096	E3	2.00
6)	Toxaphene-5	0.859	0.805	0.806	0.862	1.172	0.901	E4	17.08
7) SA	Decachlorobiphenyl	1.146	1.123	1.127	1.147	1.199	1.148	E6	2.64

Signal #2 Calibration Files

500 =PD059982.D 1000 =PD059980.D 750 =PD059981.D
 250 =PD059983.D 100 =PD059984.D

	Compound	500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.916	2.922	2.866	2.837	2.877	2.884	E6	1.23
2)	Toxaphene-1	3.138	3.111	3.123	3.091	3.239	3.141	E4	1.84
3)	Toxaphene-2	4.139	3.613	3.798	5.161	8.554	5.053	E4	40.50
4)	Toxaphene-3	7.394	7.342	7.436	7.285	7.473	7.386	E4	1.01
5)	Toxaphene-4	4.602	4.669	4.681	4.572	4.863	4.677	E4	2.42
6)	Toxaphene-5	3.352	3.327	3.430	3.335	3.343	3.357	E4	1.24
7) SA	Decachlorobiphenyl	2.347	2.284	2.311	2.358	2.442	2.348	E6	2.55

(#) = Out of Range