

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102818\
 Data File : PQ033751.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2018 15:40
 Operator : SM\SJ
 Sample : PB114247BL
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 02:11:06 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100918CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 10 08:14:42 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.560	3.839	57191826	39959216	27.879	25.692
2)	SA Decachlor...	10.383	8.882	88935579	67001247	47.048	41.679
Target Compounds							
7)	L1 AR-1016-5	0.000	5.413	0	8390	N.D.	0.184 #
8)	L2 AR-1221-1	0.000	4.028	0	21901	N.D.	1.185 #
9)	L2 AR-1221-2	4.874	4.143	315016	221741	19.135	16.764
15)	L3 AR-1232-5	0.000	5.413	0	8390	N.D.	0.430 #
20)	L4 AR-1242-5	0.000	5.727	0	10472	N.D.	0.218 #
24)	L5 AR-1248-4	0.000	5.413	0	8390	N.D.	0.122 #
25)	L5 AR-1248-5	0.000	5.850	0	14715	N.D.	0.213 #
26)	L6 AR-1254-1	0.000	5.727	0	10472	N.D.	0.093 #
27)	L6 AR-1254-2	0.000	5.850	0	14715	N.D.	0.149 #
29)	L6 AR-1254-4	7.440	6.598	14355	188242	0.116	1.769 #
30)	L6 AR-1254-5	0.000	6.995	0	8965	N.D.	0.061 #
31)	L7 AR-1260-1	7.432	0.000	66732	0	0.624	N.D. #
32)	L7 AR-1260-2	0.000	6.639	0	8343	N.D.	0.072 #
34)	L7 AR-1260-4	8.183	0.000	70676	0	0.698	N.D. #
36)	L8 AR-1262-1	0.000	6.995	0	8965	N.D.	0.070 #
43)	L9 AR-1268-3	0.000	8.050	0	41512	N.D.	0.162 #
45)	L9 AR-1268-5	0.000	8.707	0	17075	N.D.	0.020 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102818\
Data File : PQ033751.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2018 15:40
Operator : SM\SJ
Sample : PB114247BL
Misc :
ALS Vial : 54 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 02:11:06 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100918CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 10 08:14:42 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm





#7 AR-1016-5

R.T.: 0.000 min
Exp R.T.: 6.238 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :

#7 AR-1016-5


R.T.: 5.413 min
Delta R.T.: 0.003 min
Response: 8390
Conc: 0.18 ng/ml

#8 AR-1221-1

R.T.: 0.000 min
Exp R.T.: 4.792 min
Response: 0
Conc: N.D.

#8 AR-1221-1


R.T.: 4.028 min
Delta R.T.: -0.048 min
Response: 21901
Conc: 1.19 ng/ml



#9 AR-1221-2


R.T.: 4.874 min
Delta R.T.: -0.005 min
Response: 315016
Conc: 19.13 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#9 AR-1221-2

R.T.: 4.143 min
Delta R.T.: -0.020 min
Response: 221741
Conc: 16.76 ng/ml




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 6.032 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 5.413 min
Delta R.T.: 0.004 min
Response: 8390
Conc: 0.43 ng/ml



#20 AR-1242-5


R.T.: 0.000 min
Exp R.T.: 6.682 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :




#20 AR-1242-5

R.T.: 5.727 min
Delta R.T.: -0.033 min
Response: 10472
Conc: 0.22 ng/ml




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T.: 6.642 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 5.413 min
Delta R.T.: 0.005 min
Response: 8390
Conc: 0.12 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 6.680 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :




#25 AR-1248-5

R.T.: 5.850 min
Delta R.T.: 0.048 min
Response: 14715
Conc: 0.21 ng/ml




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T.: 6.615 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 5.727 min
Delta R.T.: -0.033 min
Response: 10472
Conc: 0.09 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T.: 6.833 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :




#27 AR-1254-2

R.T.: 5.850 min
Delta R.T.: -0.056 min
Response: 14715
Conc: 0.15 ng/ml




#29 AR-1254-4

R.T.: 7.440 min
Delta R.T.: -0.048 min
Response: 14355
Conc: 0.12 ng/ml



#29 AR-1254-4


R.T.: 6.598 min
Delta R.T.: 0.058 min
Response: 188242
Conc: 1.77 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T.: 7.907 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :




#30 AR-1254-5

R.T.: 6.995 min
Delta R.T.: 0.036 min
Response: 8965
Conc: 0.06 ng/ml




#31 AR-1260-1

R.T.: 7.432 min
Delta R.T.: 0.067 min
Response: 66732
Conc: 0.62 ng/ml



#31 AR-1260-1


R.T.: 0.000 min
Exp R.T.: 6.444 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T.: 7.620 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :




#32 AR-1260-2

R.T.: 6.639 min
Delta R.T.: 0.009 min
Response: 8343
Conc: 0.07 ng/ml




#34 AR-1260-4

R.T.: 8.183 min
Delta R.T.: -0.030 min
Response: 70676
Conc: 0.70 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T.: 7.257 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T.: 7.980 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :




#36 AR-1262-1

R.T.: 6.995 min
Delta R.T.: 0.000 min
Response: 8965
Conc: 0.07 ng/ml




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T.: 9.197 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 8.050 min
Delta R.T.: 0.000 min
Response: 41512
Conc: 0.16 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 10.071 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :



#45 AR-1268-5

R.T.: 8.707 min
Delta R.T.: 0.054 min
Response: 17075
Conc: 0.02 ng/ml