

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Data\PQ102818\  
 Data File : PQ033796.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 28 Oct 2018 03:18  
 Operator : SM\SJ  
 Sample : J5608-08  
 Misc :  
 ALS Vial : 96 Sample Multiplier: 1

**Instrument :**  
 ECD\_Q  
**ClientSampled :**  
 BE9K3

**Manual Integrations**  
**APPROVED**  
 Sohil  
 10/30/2018 9:18:18 AM

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Oct 29 02:21:10 2018  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Method\PQ100918CLP.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Wed Oct 10 08:14:42 2018  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
-----						
System Monitoring Compounds						
1) SA Tetrachlo...	4.561	3.838	36845853	26829670	17.961	17.250m
2) SA Decachlor...	10.380	8.879	44166049	32229933	23.365m	20.049

Target Compounds

-----  
 (f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Data\PQ102818\  
 Data File : PQ033796.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 28 Oct 2018 03:18  
 Operator : SM\SJ  
 Sample : J5608-08  
 Misc :  
 ALS Vial : 96 Sample Multiplier: 1

**Instrument :**  
 ECD\_Q  
**ClientSampled :**  
 BE9K3

**Manual Integrations**  
**APPROVED**  
 Sohil  
 10/30/2018 9:18:18 AM

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Oct 29 02:21:10 2018  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Method\PQ100918CLP.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Wed Oct 10 08:14:42 2018  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

