

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ103019\
 Data File : PQ044622.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Oct 2019 15:09
 Operator : HP\AJ
 Sample : K5111-07
 Misc : AR1248 LOD 25 PPB
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOD-MDL-WATER-01-QT4-2019

Manual Integrations
 APPROVED

Ankita
 10/31/2019 9:55:18 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 15:34:08 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ103019.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 30 15:29:39 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.307	4.339	30496973	16192917	14.087	13.312
2)	SA Decachlor...	11.546	9.761	65695433	58902854	13.843	13.798
Target Compounds							
21)	L5 AR-1248-1	6.622	5.620	989978	776883	22.755m	22.526m
22)	L5 AR-1248-2	6.917	5.881	1202425	1472703	18.141m	27.913m#
23)	L5 AR-1248-3	7.139	5.928	1567927	985737	21.018m	18.101m
24)	L5 AR-1248-4	7.565	6.113	1664335	1682010	20.409m	25.115m
25)	L5 AR-1248-5	7.608	6.539	1335605	1307012	16.122m	18.242m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ103019\
Data File : PQ044622.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Oct 2019 15:09
Operator : HP\AJ
Sample : K5111-07
Misc : AR1248 LOD 25 PPB
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT4-2019

Manual Integrations
APPROVED

Ankita
10/31/2019 9:55:18 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 15:34:08 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ103019.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 30 15:29:39 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

