

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ103118\
 Data File : PQ033944.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Oct 2018 09:15
 Operator : SM\SJ
 Sample : AR1242CCC400
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1242357

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 31 09:31:10 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100918CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 30 17:29:03 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.557	3.837	47443493	35289870	23.127	22.690
2) SA Decachlor...	10.373	8.873	63878048	61725914	33.793	38.397
Target Compounds						
16) L4 AR-1242-1	5.754	4.945	37028623	30375235	620.487	624.257
17) L4 AR-1242-2	5.754	4.977	37028623	2050694	426.361	30.381 #
18) L4 AR-1242-3	5.817	5.162	21051875	11597462	402.608	316.605
19) L4 AR-1242-4	5.916	5.236	14846415	5213212	344.272	141.536 #
20) L4 AR-1242-5	6.653	5.769	19311824	16395928	391.742	341.301

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ103118\
Data File : PQ033944.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2018 09:15
Operator : SM\SJ
Sample : AR1242CCC400
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242357

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 09:31:10 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100918CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 30 17:29:03 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

