

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ110219.M
 Title : GC EXTRACTABLES
 Last Update : Sat Nov 02 07:31:02 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ044731.D 1000 =PQ044729.D 750 =PQ044730.D
 250 =PQ044732.D 50 =PQ044733.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	7.202	7.201	7.062	7.465	8.296	7.445 E6	6.68
2) SA	Decachlorobiphenyl	6.551	6.251	6.321	6.664	7.272	6.612 E6	6.13
3) L1	AR-1016-1	2.156	2.071	2.067	2.273	2.660	2.245 E5	10.99
4) L1	AR-1016-2	3.255	3.168	3.153	3.451	3.851	3.375 E5	8.62
5) L1	AR-1016-3	2.092	2.006	2.016	2.233	2.499	2.169 E5	9.48
6) L1	AR-1016-4	1.677	1.634	1.624	1.756	1.956	1.730 E5	7.92
7) L1	AR-1016-5	1.632	1.590	1.599	1.796	2.413	1.806 E5	19.35
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1	2.682	2.559	2.591	2.781	3.154	2.753 E5	8.71
32) L7	AR-1260-2	3.049	2.935	2.962	3.237	3.939	3.224 E5	12.92
33) L7	AR-1260-3	2.378	2.291	2.332	2.501	2.806	2.462 E5	8.46
34) L7	AR-1260-4	2.644	2.561	2.607	2.720	3.131	2.733 E5	8.42
35) L7	AR-1260-5	4.954	4.858	4.873	5.015	5.746	5.089 E5	7.32
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

500 =PQ044731.D 1000 =PQ044729.D 750 =PQ044730.D
 250 =PQ044732.D 50 =PQ044733.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	3.484	3.607	3.514	3.515	3.807	3.585 E6	3.69
2) SA	Decachlorobiphenyl	7.670	7.644	8.038	7.817	8.609	7.956 E6	5.00
3) L1	AR-1016-1	1.379	1.349	1.344	1.428	1.518	1.404 E5	5.14
4) L1	AR-1016-2	2.012	2.027	1.972	2.054	2.202	2.053 E5	4.29
5) L1	AR-1016-3	1.061	1.059	1.031	1.084	1.136	1.074 E5	3.66
6) L1	AR-1016-4	0.923	0.891	0.883	0.970	1.066	0.946 E5	7.92

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ110219.M
 Title : GC EXTRACTABLES
 Last Update : Sat Nov 02 07:31:02 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ044731.D 1000 =PQ044729.D 750 =PQ044730.D
 250 =PQ044732.D 50 =PQ044733.D

Compound			500	1000	750	250	50	Avg	%RSD
7)	L1	AR-1016-5	1.208	1.154	1.151	1.264	1.469	1.249 E5	10.51
8)	L2	AR-1221-1						0.000	-1.00
9)	L2	AR-1221-2						0.000	-1.00
10)	L2	AR-1221-3						0.000	-1.00
11)	L3	AR-1232-1						0.000	-1.00
12)	L3	AR-1232-2						0.000	-1.00
13)	L3	AR-1232-3						0.000	-1.00
14)	L3	AR-1232-4						0.000	-1.00
15)	L3	AR-1232-5						0.000	-1.00
16)	L4	AR-1242-1						0.000	-1.00
17)	L4	AR-1242-2						0.000	-1.00
18)	L4	AR-1242-3						0.000	-1.00
19)	L4	AR-1242-4						0.000	-1.00
20)	L4	AR-1242-5						0.000	-1.00
21)	L5	AR-1248-1						0.000	-1.00
22)	L5	AR-1248-2						0.000	-1.00
23)	L5	AR-1248-3						0.000	-1.00
24)	L5	AR-1248-4						0.000	-1.00
25)	L5	AR-1248-5						0.000	-1.00
26)	L6	AR-1254-1						0.000	-1.00
27)	L6	AR-1254-2						0.000	-1.00
28)	L6	AR-1254-3						0.000	-1.00
29)	L6	AR-1254-4						0.000	-1.00
30)	L6	AR-1254-5						0.000	-1.00
31)	L7	AR-1260-1	2.469	2.426	2.430	2.563	2.782	2.534 E5	5.88
32)	L7	AR-1260-2	3.169	3.121	3.088	3.235	3.768	3.276 E5	8.55
33)	L7	AR-1260-3	2.916	2.905	2.872	2.925	3.115	2.946 E5	3.27
34)	L7	AR-1260-4	2.685	2.677	2.648	2.688	2.878	2.715 E5	3.41
35)	L7	AR-1260-5	6.920	7.056	6.965	6.731	6.731	6.881 E5	2.10
36)	L8	AR-1262-1						0.000	-1.00
37)	L8	AR-1262-2						0.000	-1.00
38)	L8	AR-1262-3						0.000	-1.00
39)	L8	AR-1262-4						0.000	-1.00
40)	L8	AR-1262-5						0.000	-1.00
41)	L9	AR-1268-1						0.000	-1.00
42)	L9	AR-1268-2						0.000	-1.00
43)	L9	AR-1268-3						0.000	-1.00
44)	L9	AR-1268-4						0.000	-1.00
45)	L9	AR-1268-5						0.000	-1.00

(#) = Out of Range