

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ110420\
Data File : PQ050863.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2020 11:29
Operator : DD\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
ECD_Q
LabSampleId :
AR1254334

Manual Integrations
APPROVED

Ankita
11/5/2020 11:42:16 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 03:23:21 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110320CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 04 08:52:38 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
R.T. Response Conc
7.48 575714701 853.28
7.70 591917826 571.97
8.09 654845051 583.89
8.38 332052971 413.24
8.81 361241807 416.86

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
6.40 114390597 417.06
6.56 97606609 414.68
6.98 169733379 427.21
7.22 109103251 445.88
7.65 150588358 441.35

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ110420\
Data File : PQ050863.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2020 11:29
Operator : DD\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
ECD_Q
LabSampleId :
AR1254334

Manual Integrations
APPROVED

Ankita
11/5/2020 11:42:16 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 03:23:21 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110320CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 04 08:52:38 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 #2 (L6)
R.T. Response Conc
7.48 433316808 642.23
7.70 437586077 422.84
8.09 516619246 460.64
8.38 332052971 413.24
8.81 361241807 416.86

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
6.40 114390597 417.06
6.56 97606609 414.68
6.98 169733379 427.21
7.22 109103251 445.88
7.65 150588358 441.35

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ110420\
 Data File : PQ050863.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2020 11:29
 Operator : DD\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 LabSampleId :
 AR1254334

Manual Integrations
 APPROVED

Ankita
 11/5/2020 11:42:16 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 03:23:21 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110320CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 04 08:52:38 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.241	4.258	380.4E6	97549896	21.652	21.684
2) SA Decachlor...	11.444	9.626	635.8E6	187.9E6	46.184	48.866
Target Compounds						
26) L6 AR-1254-1	7.478	6.399	433.3E6	114.4E6	642.231m	417.061 #
27) L6 AR-1254-2	7.704	6.557	437.6E6	97606609	422.840m	414.678
28) L6 AR-1254-3	8.087	6.980	516.6E6	169.7E6	460.642m	427.207
29) L6 AR-1254-4	8.382	7.218	332.1E6	109.1E6	413.244	445.878
30) L6 AR-1254-5	8.812	7.649	361.2E6	150.6E6	416.862	441.353

AJ
 11/06/20

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.