

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ110420\
Data File : PQ050866.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2020 12:18
Operator : DD\AJ
Sample : L4534-01
Misc :
ALS Vial : 9 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
C0BN7

Manual Integrations
APPROVED

Ankita
11/5/2020 11:42:19 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 03:34:16 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110320CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 04 08:52:38 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
R.T. Response Conc
7.47 204034878 302.41
7.70 383662414 370.73
8.08 488770971 435.81
8.38 420479390 523.29
8.81 733740943 846.71

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
6.40 78585120 286.52
6.56 91996815 390.84
6.98 179680699 452.24
7.22 130789722 534.51
7.65 273343100 801.13

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ110420\
Data File : PQ050866.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2020 12:18
Operator : DD\AJ
Sample : L4534-01
Misc :
ALS Vial : 9 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
C0BN7

Manual Integrations
APPROVED

Ankita
11/5/2020 11:42:19 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 03:34:16 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110320CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 04 08:52:38 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 #2 (L6)
R.T. Response Conc
7.47 204034878 302.41
7.70 383662414 370.73
8.08 488770971 435.81
8.38 420479390 523.29
8.81 636187469 734.14
(26) AR-1254-1 #2 (L6)
R.T. Response Conc
6.40 78585120 286.52
6.56 91996815 390.84
6.98 179680699 452.24
7.22 130789722 534.51
7.65 273343100 801.13

(+) = Expected Retention Time

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ110420\
 Data File : PQ050866.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2020 12:18
 Operator : DD\AJ
 Sample : L4534-01
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 C0BN7

Manual Integrations
 APPROVED

Ankita
 11/5/2020 11:42:19 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 03:34:16 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110320CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 04 08:52:38 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.237	4.259	298.1E6	74359986	16.968	16.529
2) SA Decachlor...	11.436	9.624	436.3E6	123.8E6	31.689	32.185
Target Compounds						
26) L6 AR-1254-1	7.473	6.399	204.0E6	78585120	302.406	286.516
27) L6 AR-1254-2	7.700	6.557	383.7E6	91996815	370.733	390.845
28) L6 AR-1254-3	8.085	6.979	488.8E6	179.7E6	435.811	452.244
29) L6 AR-1254-4	8.379	7.217	420.5E6	130.8E6	523.291	534.505
30) L6 AR-1254-5	8.807	7.648	636.2E6	273.3E6	734.141m	801.130

AJ
 11/06/20

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.