

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ110519.M
 Title : GC EXTRACTABLES
 Last Update : Wed Nov 06 05:33:47 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ044771.D 1000 =PQ044769.D 750 =PQ044770.D
 250 =PQ044772.D 50 =PQ044773.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	5.264	4.824	5.240	5.530	5.290	5.229	E6	4.87
2) SA	Decachlorobiphenyl	1.112	1.059	1.098	1.122	1.163	1.111	E7	3.43
3) L1	AR-1016-1	1.643	1.476	1.624	1.783	1.764	1.658	E5	7.47
4) L1	AR-1016-2	2.598	2.340	2.523	2.736	2.704	2.580	E5	6.15
5) L1	AR-1016-3	1.576	1.403	1.530	1.671	1.658	1.567	E5	6.94
6) L1	AR-1016-4	1.294	1.160	1.261	1.355	1.373	1.288	E5	6.60
7) L1	AR-1016-5	1.417	1.251	1.366	1.489	1.497	1.404	E5	7.21
8) L2	AR-1221-1	4.206					4.206	E4	0.00
9) L2	AR-1221-2	3.325					3.325	E4	0.00
10) L2	AR-1221-3	1.099					1.099	E5	0.00
11) L3	AR-1232-1	1.513					1.513	E5	0.00
12) L3	AR-1232-2	8.572					8.572	E4	0.00
13) L3	AR-1232-3	1.635					1.635	E5	0.00
14) L3	AR-1232-4	8.030					8.030	E4	0.00
15) L3	AR-1232-5	6.354					6.354	E4	0.00
16) L4	AR-1242-1	1.100					1.100	E5	0.00
17) L4	AR-1242-2	1.715					1.715	E5	0.00
18) L4	AR-1242-3	1.050					1.050	E5	0.00
19) L4	AR-1242-4	8.532					8.532	E4	0.00
20) L4	AR-1242-5	1.087					1.087	E5	0.00
21) L5	AR-1248-1	9.567					9.567	E4	0.00
22) L5	AR-1248-2	1.502					1.502	E5	0.00
23) L5	AR-1248-3	1.745					1.745	E5	0.00
24) L5	AR-1248-4	1.991					1.991	E5	0.00
25) L5	AR-1248-5	1.959					1.959	E5	0.00
26) L6	AR-1254-1	2.160					2.160	E5	0.00
27) L6	AR-1254-2	3.495					3.495	E5	0.00
28) L6	AR-1254-3	4.117					4.117	E5	0.00
29) L6	AR-1254-4	3.143					3.143	E5	0.00
30) L6	AR-1254-5	4.600					4.600	E5	0.00
31) L7	AR-1260-1	3.280	2.990	3.206	3.429	3.587	3.298	E5	6.85
32) L7	AR-1260-2	4.059	3.701	3.969	4.185	4.393	4.061	E5	6.32
33) L7	AR-1260-3	3.606	3.297	3.525	3.682	3.851	3.592	E5	5.69
34) L7	AR-1260-4	4.394	4.064	4.293	4.475	4.866	4.418	E5	6.66
35) L7	AR-1260-5	8.997	8.498	8.924	8.961	9.177	8.911	E5	2.81
36) L8	AR-1262-1	5.235					5.235	E5	0.00
37) L8	AR-1262-2	1.016					1.016	E6	0.00
38) L8	AR-1262-3	6.915					6.915	E5	0.00
39) L8	AR-1262-4	5.254					5.254	E5	0.00
40) L8	AR-1262-5	3.653					3.653	E5	0.00
41) L9	AR-1268-1	1.211					1.211	E6	0.00
42) L9	AR-1268-2	1.113					1.113	E6	0.00
43) L9	AR-1268-3	9.884					9.884	E5	0.00
44) L9	AR-1268-4	3.844					3.844	E5	0.00
45) L9	AR-1268-5	3.495					3.495	E6	0.00

Signal #2 Calibration Files

500 =PQ044771.D 1000 =PQ044769.D 750 =PQ044770.D
 250 =PQ044772.D 50 =PQ044773.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.149	3.821	4.155	4.293	3.905	4.065	E6	4.80
2) SA	Decachlorobiphenyl	8.742	8.247	8.551	8.720	9.292	8.710	E6	4.37
3) L1	AR-1016-1	1.558	1.437	1.509	1.682	1.731	1.583	E5	7.67
4) L1	AR-1016-2	2.268	2.055	2.233	2.376	2.452	2.277	E5	6.65
5) L1	AR-1016-3	1.197	1.098	1.165	1.259	1.328	1.209	E5	7.31

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ110519.M
 Title : GC EXTRACTABLES
 Last Update : Wed Nov 06 05:33:47 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ044771.D 1000 =PQ044769.D 750 =PQ044770.D
 250 =PQ044772.D 50 =PQ044773.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	1.027	0.919	0.984	1.104	1.191	1.045	E5	10.11
7)	L1 AR-1016-5	1.348	1.220	1.300	1.441	1.479	1.358	E5	7.71
8)	L2 AR-1221-1	3.636					3.636	E4	0.00
9)	L2 AR-1221-2	2.900					2.900	E4	0.00
10)	L2 AR-1221-3	9.601					9.601	E4	0.00
11)	L3 AR-1232-1	1.316					1.316	E5	0.00
12)	L3 AR-1232-2	1.484					1.484	E5	0.00
13)	L3 AR-1232-3	7.650					7.650	E4	0.00
14)	L3 AR-1232-4	7.160					7.160	E4	0.00
15)	L3 AR-1232-5	7.948					7.948	E4	0.00
16)	L4 AR-1242-1	1.080					1.080	E5	0.00
17)	L4 AR-1242-2	1.561					1.561	E5	0.00
18)	L4 AR-1242-3	8.261					8.261	E4	0.00
19)	L4 AR-1242-4	8.596					8.596	E4	0.00
20)	L4 AR-1242-5	1.158					1.158	E5	0.00
21)	L5 AR-1248-1	9.331					9.331	E4	0.00
22)	L5 AR-1248-2	1.318					1.318	E5	0.00
23)	L5 AR-1248-3	1.353					1.353	E5	0.00
24)	L5 AR-1248-4	1.656					1.656	E5	0.00
25)	L5 AR-1248-5	1.664					1.664	E5	0.00
26)	L6 AR-1254-1	2.573					2.573	E5	0.00
27)	L6 AR-1254-2	2.280					2.280	E5	0.00
28)	L6 AR-1254-3	3.891					3.891	E5	0.00
29)	L6 AR-1254-4	2.585					2.585	E5	0.00
30)	L6 AR-1254-5	3.762					3.762	E5	0.00
31)	L7 AR-1260-1	2.907	2.724	2.767	3.013	3.124	2.907	E5	5.75
32)	L7 AR-1260-2	3.331	3.414	3.490	3.581	3.607	3.484	E5	3.30
33)	L7 AR-1260-3	3.477	3.229	3.312	3.438	3.405	3.372	E5	2.99
34)	L7 AR-1260-4	3.189	2.948	3.070	3.092	3.171	3.094	E5	3.10
35)	L7 AR-1260-5	8.144	7.740	7.985	7.623	7.585	7.815	E5	3.08
36)	L8 AR-1262-1	2.057					2.057	E5	0.00
37)	L8 AR-1262-2	8.493					8.493	E5	0.00
38)	L8 AR-1262-3	3.340					3.340	E5	0.00
39)	L8 AR-1262-4	6.406					6.406	E5	0.00
40)	L8 AR-1262-5	3.043					3.043	E5	0.00
41)	L9 AR-1268-1	9.561					9.561	E5	0.00
42)	L9 AR-1268-2	8.855					8.855	E5	0.00
43)	L9 AR-1268-3	7.439					7.439	E5	0.00
44)	L9 AR-1268-4	3.273					3.273	E5	0.00
45)	L9 AR-1268-5	2.630					2.630	E6	0.00

(#) = Out of Range