

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ110525\
 Data File : PQ071200.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Nov 2025 10:59
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 12:10:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110525.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 11:28:10 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.411	2.860	438.3E6	534.9E6	74.397	74.381
2)	SA Decachlor...	8.538	7.669	295.1E6	480.8E6	73.918	73.780
Target Compounds							
16)	L4 AR-1242-1	4.497	3.876	129.1E6	164.2E6	745.242	742.729
17)	L4 AR-1242-2	4.516	3.893	189.0E6	229.5E6	741.440	742.956
18)	L4 AR-1242-3	4.574	4.055	114.5E6	124.9E6	744.694	740.719
19)	L4 AR-1242-4	4.664	4.142	94735079	111.0E6	746.881	742.364
20)	L4 AR-1242-5	5.338	4.641	92386284	150.8E6	744.639	737.756

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ110525\
Data File : PQ071200.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Nov 2025 10:59
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 12:10:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110525.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 11:28:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

