

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ110719\
 Data File : PQ044894.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Nov 2019 09:42
 Operator : SM\AJ
 Sample : K5678-19
 Misc :
 ALS Vial : 63 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 JLMK4

Manual Integrations
 APPROVED

Ankita
 11/11/2019 3:27:33 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 23:10:56 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110719CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 08 03:35:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.102	4.309	102.1E6	59824924	12.865	12.688
2)	SA Decachlor...	11.170	9.726	159.8E6	157.2E6	23.066	23.168
Target Compounds							
31)	L7 AR-1260-1	8.106	7.192	24876900	26984936	79.651	88.079
32)	L7 AR-1260-2	8.369	7.387	44042775	36439005	107.569	92.284
33)	L7 AR-1260-3	8.737	7.549	36301210	32356743	111.190	94.448
34)	L7 AR-1260-4	8.969	8.034	39524511	31540940	103.938m	102.854
35)	L7 AR-1260-5	9.305	8.280	70896334	75373438	92.097	96.308

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ110719\
Data File : PQ044894.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Nov 2019 09:42
Operator : SM\AJ
Sample : K5678-19
Misc :
ALS Vial : 63 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
JLMK4

Manual Integrations
APPROVED

Ankita
11/11/2019 3:27:33 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 08 23:10:56 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110719CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 08 03:35:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

