

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ110818\
Data File : PQ034318.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Nov 2018 06:21
Operator : SM\SJ
Sample : J5851-15
Misc :
ALS Vial : 76 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
110618-SIDI-F-U-S

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 10 00:38:58 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110618.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 09 08:12:53 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.549	3.829	48607105	34490066	16.582	18.070
2)	SA Decachlor...	10.353	8.861	40099021	31766868	15.621	16.165

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ110818\
Data File : PQ034318.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Nov 2018 06:21
Operator : SM\SJ
Sample : J5851-15
Misc :
ALS Vial : 76 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
110618-SIDI-F-U-S

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 10 00:38:58 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110618.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 09 08:12:53 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.549 min
Delta R.T.: -0.005 min
Response: 48607105
Conc: 16.58 ng/ml

Instrument :
ECD_Q
ClientSampleId :
110618-SIDI-F-U-S




#1 Tetrachloro-m-xylene

R.T.: 3.829 min
Delta R.T.: -0.003 min
Response: 34490066
Conc: 18.07 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.353 min
Delta R.T.: -0.012 min
Response: 40099021
Conc: 15.62 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.861 min
Delta R.T.: -0.006 min
Response: 31766868
Conc: 16.17 ng/ml