

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ111023\
 Data File : PQ063914.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Nov 2023 09:55
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 144 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 10 22:27:44 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 08 18:43:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.460	2.777	224.2E6	211.0E6	52.588	57.379
2) SA Decachlor...	8.695	7.575	160.2E6	252.1E6	49.615	53.813
Target Compounds						
26) L6 AR-1254-1	5.405	4.548	83234512	100.2E6	530.326	515.745
27) L6 AR-1254-2	5.624	4.696	126.6E6	89647480	533.899	514.611
28) L6 AR-1254-3	5.980	5.080	130.5E6	138.7E6	526.192	518.003
29) L6 AR-1254-4	6.267	5.307	95050497	90444518	524.496	515.559
30) L6 AR-1254-5	6.682	5.714	107.0E6	133.0E6	536.321	521.881

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ111023\
Data File : PQ063914.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Nov 2023 09:55
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 144 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 10 22:27:44 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110823.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 08 18:43:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

