

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ111123\
 Data File : PQ063950.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Nov 2023 20:55
 Operator : YP\AJ
 Sample : AR1242ICC400
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 11 00:48:08 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111123CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Nov 11 00:45:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.460	2.777	94920659	93565616	20.000	20.000
2) SA Decachlor...	8.693	7.573	137.8E6	223.3E6	40.000	40.000
Target Compounds						
16) L4 AR-1242-1	4.574	3.787	50016830	46764760	400.000	400.000
17) L4 AR-1242-2	4.593	3.802	74611405	67756900	400.000	400.000
18) L4 AR-1242-3	4.652	3.964	46500676	36868901	400.000	400.000
19) L4 AR-1242-4	4.744	4.050	37692060	34024582	400.000	400.000
20) L4 AR-1242-5	5.466	4.548	40074336	43293914	400.000	400.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ111123\
Data File : PQ063950.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Nov 2023 20:55
Operator : YP\AJ
Sample : AR1242ICC400
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 11 00:48:08 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111123CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Nov 11 00:45:16 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.460 min
Delta R.T.: 0.000 min
Response: 94920659
Conc: 20.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.777 min
Delta R.T.: 0.000 min
Response: 93565616
Conc: 20.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.693 min
Delta R.T.: 0.000 min
Response: 137786054
Conc: 40.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.573 min
Delta R.T.: 0.000 min
Response: 223329859
Conc: 40.00 ng/ml



#16 AR-1242-1


R.T.: 4.574 min
Delta R.T.: 0.000 min
Response: 50016830
Conc: 400.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#16 AR-1242-1

R.T.: 3.787 min
Delta R.T.: 0.000 min
Response: 46764760
Conc: 400.00 ng/ml




#17 AR-1242-2

R.T.: 4.593 min
Delta R.T.: 0.000 min
Response: 74611405
Conc: 400.00 ng/ml



#17 AR-1242-2


R.T.: 3.802 min
Delta R.T.: 0.000 min
Response: 67756900
Conc: 400.00 ng/ml



#18 AR-1242-3


R.T.: 4.652 min
Delta R.T.: 0.000 min
Response: 46500676
Conc: 400.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#18 AR-1242-3

R.T.: 3.964 min
Delta R.T.: 0.000 min
Response: 36868901
Conc: 400.00 ng/ml




#19 AR-1242-4

R.T.: 4.744 min
Delta R.T.: 0.000 min
Response: 37692060
Conc: 400.00 ng/ml



#19 AR-1242-4


R.T.: 4.050 min
Delta R.T.: 0.000 min
Response: 34024582
Conc: 400.00 ng/ml



#20 AR-1242-5

R.T.: 5.466 min
Delta R.T.: 0.000 min
Response: 40074336
Conc: 400.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :



#20 AR-1242-5

R.T.: 4.548 min
Delta R.T.: 0.000 min
Response: 43293914
Conc: 400.00 ng/ml