

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ111123\
 Data File : PQ063971.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Nov 2023 02:04
 Operator : YP\AJ
 Sample : AR1268ICC800
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12684027

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 11 05:13:59 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111123CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Nov 11 05:13:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.460	2.777	176.2E6	179.2E6	37.623	37.563
2)	SA Decachlor...	8.690	7.572	440.9E6	730.8E6	72.814	73.307
Target Compounds							
41)	L9 AR-1268-1	7.566	6.531	406.0E6	470.3E6	736.770	745.925
42)	L9 AR-1268-2	7.645	6.594	368.3E6	430.5E6	739.655	754.204
43)	L9 AR-1268-3	7.826	6.794	316.8E6	378.1E6	737.479	753.004
44)	L9 AR-1268-4	8.151	7.086	130.7E6	164.5E6	728.185	750.840
45)	L9 AR-1268-5	8.452	7.355	949.1E6	1412.0E6	755.923	759.190

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ111123\
Data File : PQ063971.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Nov 2023 02:04
Operator : YP\AJ
Sample : AR1268ICC800
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12684027

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 11 05:13:59 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111123CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Nov 11 05:13:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

