

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ111123\
 Data File : PQ063975.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Nov 2023 03:18
 Operator : YP\AJ
 Sample : AR1242CCC400
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12423191

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 11 06:19:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111123CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Nov 11 06:14:57 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.460	2.777	88005294	86892729	19.204	19.353
2) SA Decachlor...	8.693	7.573	127.0E6	208.2E6	38.445	38.837
Target Compounds						
16) L4 AR-1242-1	4.574	3.787	46352196	42750459	386.416	373.344
17) L4 AR-1242-2	4.594	3.802	68444378	63063586	380.001	379.098
18) L4 AR-1242-3	4.653	3.964	42883424	33628818	382.140	381.371
19) L4 AR-1242-4	4.745	4.050	35062944	31382102	380.917	381.475
20) L4 AR-1242-5	5.466	4.548	37642712	40442154	393.829	380.313

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ111123\
Data File : PQ063975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Nov 2023 03:18
Operator : YP\AJ
Sample : AR1242CCC400
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12423191

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 11 06:19:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111123CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Nov 11 06:14:57 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

