

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ111123\
 Data File : PQ063937.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Nov 2023 17:45
 Operator : YP\AJ
 Sample : AR1660ICC1600
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR16605043

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 10 17:57:28 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111123CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 10 17:57:14 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.459	2.777	344.4E6	331.2E6	74.392	74.289
2) SA Decachlor...	8.693	7.574	458.5E6	749.6E6	142.813	143.936
Target Compounds						
3) L1 AR-1016-1	4.574	3.787	204.1E6	184.8E6	1426.785	1418.536
4) L1 AR-1016-2	4.594	3.803	316.8E6	276.0E6	1461.254	1435.426
5) L1 AR-1016-3	4.652	3.964	191.8E6	147.4E6	1425.784	1435.194
6) L1 AR-1016-4	4.744	4.013	159.6E6	119.5E6	1446.057	1383.885
7) L1 AR-1016-5	5.035	4.209	157.4E6	147.7E6	1425.346	1405.695
31) L7 AR-1260-1	6.145	5.209	302.5E6	286.8E6	1425.406	1420.992
32) L7 AR-1260-2	6.407	5.399	360.1E6	347.0E6	1421.746	1436.447
33) L7 AR-1260-3	6.762	5.542	266.0E6	333.9E6	1424.431	1445.869
34) L7 AR-1260-4	6.981	6.006	296.8E6	289.1E6	1430.025	1453.321
35) L7 AR-1260-5	7.293	6.252	585.6E6	691.3E6	1479.235	1510.247

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ111123\
Data File : PQ063937.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Nov 2023 17:45
Operator : YP\AJ
Sample : AR1660ICC1600
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR16605043

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 10 17:57:28 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111123CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 10 17:57:14 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

