

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ111622\
 Data File : PQ060081.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Nov 2022 09:58
 Operator : YP\AJ
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR16603052

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 16 16:46:18 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110922CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 10 08:42:03 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.457	2.805	270.6E6	207.5E6	22.364	17.077
2)	SA Decachlor...	8.633	7.572	150.7E6	367.0E6	40.257	36.408
Target Compounds							
3)	L1 AR-1016-1	4.556	3.810	166.5E6	134.1E6	460.444	346.805
4)	L1 AR-1016-2	4.575	3.826	244.4E6	199.6E6	446.514	343.178
5)	L1 AR-1016-3	4.633	3.987	153.6E6	107.6E6	450.669	359.000
6)	L1 AR-1016-4	4.725	4.034	126.3E6	83513091	451.280	389.500
7)	L1 AR-1016-5	5.011	4.230	122.5E6	109.6E6	445.199	361.247
31)	L7 AR-1260-1	6.112	5.223	193.2E6	223.0E6	440.684	367.906
32)	L7 AR-1260-2	6.371	5.411	217.7E6	267.6E6	437.910	364.527
33)	L7 AR-1260-3	6.725	5.553	145.6E6	248.1E6	428.158	348.687
34)	L7 AR-1260-4	6.943	6.014	155.1E6	213.4E6	444.814	355.678
35)	L7 AR-1260-5	7.252	6.258	275.5E6	494.8E6	431.515	349.926

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ111622\
Data File : PQ060081.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Nov 2022 09:58
Operator : YP\AJ
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR16603052

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 16 16:46:18 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110922CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 10 08:42:03 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

