

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ111718CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 16 02:50:22 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ034629.D 1600 =PQ034631.D 800 =PQ034630.D
 200 =PQ034628.D 100 =PQ034627.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	2.841	2.705	2.729	2.786	2.902	2.793	E6	2.89
2)	SA	Decachlorobiphenyl	2.644	2.454	2.426	2.787	3.013	2.665	E6	9.16
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1	5.373	4.722	5.046	5.796	6.069	5.401	E4	10.10
12)	L3	AR-1232-2	2.826	2.582	2.683	3.073	3.237	2.880	E4	9.43
13)	L3	AR-1232-3	5.759	5.255	5.556	6.388	6.430	5.877	E4	8.80
14)	L3	AR-1232-4	2.673	2.509	2.644	2.992	3.061	2.776	E4	8.58
15)	L3	AR-1232-5	2.133	1.839	2.011	2.454	2.521	2.192	E4	13.26
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ034629.D 1600 =PQ034631.D 800 =PQ034630.D
 200 =PQ034628.D 100 =PQ034627.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.633	1.548	1.564	1.638	1.766	1.630	E6	5.28
2) SA	Decachlorobiphenyl	1.850	1.745	1.745	1.986	2.196	1.904	E6	10.02
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ111718CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 16 02:50:22 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ034629.D 1600 =PQ034631.D 800 =PQ034630.D
 200 =PQ034628.D 100 =PQ034627.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1	3.583	3.031	3.294	3.697	3.930	3.507 E4	10.01
12) L3	AR-1232-2	3.859	3.523	3.711	4.226	4.056	3.875 E4	7.15
13) L3	AR-1232-3	1.996	1.820	1.936	2.124	2.157	2.006 E4	6.88
14) L3	AR-1232-4	1.800	1.576	1.713	1.950	1.649	1.738 E4	8.32
15) L3	AR-1232-5	1.917	1.752	1.874	2.125	2.010	1.936 E4	7.28
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range