

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ111718CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 16 01:16:05 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ034644.D 1600 =PQ034646.D 800 =PQ034645.D
 200 =PQ034643.D 100 =PQ034642.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	2.546	2.844	2.621	2.704	2.780	2.699	E6	4.43
2)	SA	Decachlorobiphenyl	2.636	2.552	2.438	2.808	2.971	2.681	E6	7.86
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1	1.073	1.051	1.003	1.157	1.311	1.119	E5	10.80
27)	L6	AR-1254-2	1.700	1.669	1.583	1.820	2.074	1.769	E5	10.76
28)	L6	AR-1254-3	1.867	1.870	1.753	1.991	2.209	1.938	E5	8.93
29)	L6	AR-1254-4	1.386	1.354	1.274	1.457	1.627	1.419	E5	9.38
30)	L6	AR-1254-5	1.480	1.461	1.379	1.576	1.630	1.505	E5	6.56
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ034644.D 1600 =PQ034646.D 800 =PQ034645.D
 200 =PQ034643.D 100 =PQ034642.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.596	1.721	1.601	1.706	1.794	1.683	E6	5.02
2) SA	Decachlorobiphenyl	1.854	1.783	1.713	1.998	2.201	1.910	E6	10.17
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ111718CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 16 01:16:05 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ034644.D 1600 =PQ034646.D 800 =PQ034645.D
 200 =PQ034643.D 100 =PQ034642.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	1.045	1.030	0.980	1.106	1.303	1.093 E5	11.51
27) L6	AR-1254-2	0.908	0.881	0.854	0.977	1.159	0.956 E5	12.79
28) L6	AR-1254-3	1.550	1.541	1.462	1.653	1.878	1.617 E5	9.96
29) L6	AR-1254-4	0.996	0.977	0.928	1.055	1.219	1.035 E5	10.85
30) L6	AR-1254-5	1.374	1.342	1.278	1.448	1.652	1.419 E5	10.13
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range