

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ111718CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 16 03:29:42 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ034654.D 1600 =PQ034656.D 800 =PQ034655.D
 200 =PQ034653.D 100 =PQ034652.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.885	3.040	2.828	2.937	3.035	2.945	E6	3.15
2) SA	Decachlorobiphenyl	4.893	4.951	4.666	5.108	5.678	5.059	E6	7.52
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1	4.181	4.305	4.027	4.342	4.691	4.309	E5	5.72
42) L9	AR-1268-2	3.766	3.905	3.632	3.901	4.205	3.882	E5	5.48
43) L9	AR-1268-3	3.226	3.337	3.106	3.332	3.610	3.322	E5	5.61
44) L9	AR-1268-4	1.350	1.337	1.273	1.409	1.536	1.381	E5	7.19
45) L9	AR-1268-5	1.069	1.137	1.057	1.084	1.164	1.102	E6	4.18

Signal #2 Calibration Files

400 =PQ034654.D 1600 =PQ034656.D 800 =PQ034655.D
 200 =PQ034653.D 100 =PQ034652.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.812	1.847	1.750	1.884	2.011	1.861	E6	5.22
2) SA	Decachlorobiphenyl	3.475	3.454	3.282	3.646	3.995	3.570	E6	7.57
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ111718CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 16 03:29:42 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ034654.D 1600 =PQ034656.D 800 =PQ034655.D
 200 =PQ034653.D 100 =PQ034652.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	3.274	3.246	3.102	3.409	3.619	3.330 E5	5.85
42) L9	AR-1268-2	2.924	2.952	2.783	3.031	3.243	2.987 E5	5.66
43) L9	AR-1268-3	2.412	2.451	2.305	2.512	2.716	2.479 E5	6.14
44) L9	AR-1268-4	1.040	1.005	0.967	1.100	1.206	1.064 E5	8.77
45) L9	AR-1268-5	7.826	7.942	7.542	7.967	8.497	7.955 E5	4.36

(#) = Out of Range