Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ111720\
Data File : PQ050994.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2020 16:52
Operator : DD\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254342

Manual Integrations APPROVED
Ankita
11/18/2020 10:35:51 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 06:03:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110320CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 04 08:52:38 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	5.237	4.256	395.4E6	108.4E6	22.508	24.104
2)	SA Decachlor...	11.438	9.619	431.5E6	144.2E6	31.341	37.496
	Target Compounds						
26)	L6 AR-1254-1	7.472	6.395	211.3E6	109.4E6	313.194m	398.731 #
27)	L6 AR-1254-2	7.699	6.553	272.8E6	93602898	263.601m	397.668 #
28)	L6 AR-1254-3	8.084	6.976	337.8E6	153.7E6	301.229m	386.824 #
29)	L6 AR-1254-4	8.379	7.213	233.7E6	93678690	290.789	382.842 #
30)	L6 AR-1254-5	8.808	7.645	242.1E6	119.0E6	279.407	348.893

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ111720\
Data File : PQ050994.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Nov 2020 16:52
Operator : DD\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254342

Manual Integrations APPROVED
Ankita
11/18/2020 10:35:51 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 18 06:03:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110320CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 04 08:52:38 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm