

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ111820\
 Data File : PQ051011.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Nov 2020 18:58
 Operator : DD\AJ
 Sample : AR1232ICC1600
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1232501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 19 03:41:09 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111820CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 19 03:39:10 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.238	4.259	1616.1E6	444.8E6	70.474	75.364
2) SA Decachlor...	11.435	9.618	2435.2E6	767.5E6	145.912	150.723
Target Compounds						
11) L3 AR-1232-1	5.674	4.707	579.5E6	167.7E6	1314.181	1366.133
12) L3 AR-1232-2	6.266	5.541	321.6E6	189.1E6	1379.827	1483.330
13) L3 AR-1232-3	6.584	5.734	680.9E6	94512727	1422.690	1454.315
14) L3 AR-1232-4	6.759	5.830	338.6E6	80519049	1451.528	1411.065
15) L3 AR-1232-5	6.853	6.015	235.2E6	88220431	1426.344	1456.666

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ111820\
Data File : PQ051011.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Nov 2020 18:58
Operator : DD\AJ
Sample : AR1232ICC1600
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1232501

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 19 03:41:09 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111820CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 19 03:39:10 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

