

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ111820\
 Data File : PQ051031.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Nov 2020 00:26
 Operator : DD\AJ
 Sample : AR1262ICC1600
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1262501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 19 05:12:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111820CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 19 05:10:35 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.235	4.255	2051.1E6	546.3E6	81.999	83.136
2) SA Decachlor...	11.430	9.613	2674.6E6	834.8E6	147.756	149.540
Target Compounds						
36) L8 AR-1262-1	8.877	7.640	1933.2E6	348.7E6	1535.361	1516.571
37) L8 AR-1262-2	9.467	8.192	3545.1E6	1301.3E6	1570.351	1503.623
38) L8 AR-1262-3	9.800	8.480	1446.3E6	471.0E6	1557.310	1497.618
39) L8 AR-1262-4	9.901	8.544	872.4E6	936.0E6	1433.625	1518.684
40) L8 AR-1262-5	10.617	9.047	1223.0E6	420.3E6	1516.104	1503.324

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ11820\
Data File : PQ051031.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Nov 2020 00:26
Operator : DD\AJ
Sample : AR1262ICC1600
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1262501

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 19 05:12:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ11820CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 19 05:10:35 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

