

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112020\
 Data File : PQ051158.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2020 17:34
 Operator : DD\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254347

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 00:38:50 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111820CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 19 05:59:02 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.236	4.256	496.1E6	129.9E6	21.909	21.792
2)	SA Decachlor...	11.433	9.612	708.3E6	209.3E6	43.168	40.988
Target Compounds							
26)	L6 AR-1254-1	7.470	6.392	339.8E6	136.2E6	396.265	380.334
27)	L6 AR-1254-2	7.698	6.549	524.4E6	116.7E6	393.471	376.604
28)	L6 AR-1254-3	8.082	6.971	572.1E6	194.8E6	395.018	376.916
29)	L6 AR-1254-4	8.376	7.209	415.1E6	121.9E6	396.303	374.360
30)	L6 AR-1254-5	8.805	7.640	438.2E6	169.1E6	397.239	372.623

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112020\
Data File : PQ051158.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2020 17:34
Operator : DD\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
AR1254347

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 00:38:50 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111820CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 19 05:59:02 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

