

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112023\
 Data File : PQ064201.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2023 19:23
 Operator : YP\AJ
 Sample : AR1232ICC800
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 00:13:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ112023CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 00:09:37 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
1)	SA Tetrachlo...	3.457	2.775	175.1E6	178.7E6	37.874	37.502
2)	SA Decachlor...	8.688	7.571	228.7E6	281.5E6	75.342	75.126
Target Compounds							
11)	L3 AR-1232-1	3.811	3.124	70295523	76540888	717.121	702.943
12)	L3 AR-1232-2	4.311	3.800	36704845	68794051	719.963	738.232
13)	L3 AR-1232-3	4.591	3.961	68390247	34803522	720.172	729.724
14)	L3 AR-1232-4	4.742	4.047	35054164	29899380	737.424	729.008
15)	L3 AR-1232-5	4.840	4.206	25734012	34120032	731.040	727.064


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112023\
Data File : PQ064201.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2023 19:23
Operator : YP\AJ
Sample : AR1232ICC800
Misc :
ALS Vial : 16 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 00:13:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ112023CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 00:09:37 2023
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.457 min
Delta R.T.: 0.000 min
Response: 175107256
Conc: 37.87 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 2.775 min
Delta R.T.: 0.000 min
Response: 178667741
Conc: 37.50 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.688 min
Delta R.T.: 0.000 min
Response: 228723628
Conc: 75.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 7.571 min
Delta R.T.: 0.001 min
Response: 281450611
Conc: 75.13 ng/ml




#11 AR-1232-1

R.T.: 3.811 min
Delta R.T.: 0.000 min
Response: 70295523
Conc: 717.12 ng/ml




#11 AR-1232-1

R.T.: 3.124 min
Delta R.T.: 0.000 min
Response: 76540888
Conc: 702.94 ng/ml




#12 AR-1232-2

R.T.: 4.311 min
Delta R.T.: 0.000 min
Response: 36704845
Conc: 719.96 ng/ml




#12 AR-1232-2

R.T.: 3.800 min
Delta R.T.: 0.000 min
Response: 68794051
Conc: 738.23 ng/ml




#13 AR-1232-3

R.T.: 4.591 min
Delta R.T.: 0.000 min
Response: 68390247
Conc: 720.17 ng/ml




#13 AR-1232-3

R.T.: 3.961 min
Delta R.T.: 0.000 min
Response: 34803522
Conc: 729.72 ng/ml




#14 AR-1232-4

R.T.: 4.742 min
Delta R.T.: 0.000 min
Response: 35054164
Conc: 737.42 ng/ml




#14 AR-1232-4

R.T.: 4.047 min
Delta R.T.: 0.000 min
Response: 29899380
Conc: 729.01 ng/ml



#15 AR-1232-5

R.T.: 4.840 min
Delta R.T.: 0.000 min
Response: 25734012
Conc: 731.04 ng/ml



#15 AR-1232-5

R.T.: 4.206 min
Delta R.T.: 0.000 min
Response: 34120032
Conc: 727.06 ng/ml