

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112023\
 Data File : PQ064206.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2023 20:36
 Operator : YP\AJ
 Sample : AR1242ICC800
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 00:30:13 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ112023CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 00:28:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds						
1) SA Tetrachlo...	3.457	2.774	181.8E6	191.1E6	39.169	39.098
2) SA Decachlor...	8.689	7.569	239.4E6	295.5E6	77.136	76.804
Target Compounds						
16) L4 AR-1242-1	4.572	3.784	90083507	87319897	776.392	750.143
17) L4 AR-1242-2	4.591	3.799	137.7E6	130.8E6	776.688	762.348
18) L4 AR-1242-3	4.650	3.961	86032310	68388724	773.949	762.427
19) L4 AR-1242-4	4.742	4.047	70396092	63350989	774.835	746.553
20) L4 AR-1242-5	5.465	4.545	74595148	80037034	801.188	762.161


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112023\
Data File : PQ064206.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2023 20:36
Operator : YP\AJ
Sample : AR1242ICC800
Misc :
ALS Vial : 21 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 00:30:13 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ112023CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 00:28:50 2023
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.457 min
Delta R.T.: 0.000 min
Response: 181781630
Conc: 39.17 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: 0.000 min
Response: 191140711
Conc: 39.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.689 min
Delta R.T.: 0.000 min
Response: 239356916
Conc: 77.14 ng/ml




#2 Decachlorobiphenyl

R.T.: 7.569 min
Delta R.T.: 0.000 min
Response: 295477330
Conc: 76.80 ng/ml




#16 AR-1242-1

R.T.: 4.572 min
Delta R.T.: 0.000 min
Response: 90083507
Conc: 776.39 ng/ml




#16 AR-1242-1

R.T.: 3.784 min
Delta R.T.: 0.000 min
Response: 87319897
Conc: 750.14 ng/ml




#17 AR-1242-2

R.T.: 4.591 min
Delta R.T.: 0.000 min
Response: 137651044
Conc: 776.69 ng/ml




#17 AR-1242-2

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 130790505
Conc: 762.35 ng/ml




#18 AR-1242-3

R.T.: 4.650 min
Delta R.T.: 0.000 min
Response: 86032310
Conc: 773.95 ng/ml




#18 AR-1242-3

R.T.: 3.961 min
Delta R.T.: 0.000 min
Response: 68388724
Conc: 762.43 ng/ml




#19 AR-1242-4

R.T.: 4.742 min
Delta R.T.: 0.000 min
Response: 70396092
Conc: 774.84 ng/ml




#19 AR-1242-4

R.T.: 4.047 min
Delta R.T.: 0.000 min
Response: 63350989
Conc: 746.55 ng/ml



#20 AR-1242-5

R.T.: 5.465 min
Delta R.T.: 0.002 min
Response: 74595148
Conc: 801.19 ng/ml



#20 AR-1242-5

R.T.: 4.545 min
Delta R.T.: 0.000 min
Response: 80037034
Conc: 762.16 ng/ml