

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112023\
 Data File : PQ064225.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2023 01:15
 Operator : YP\AJ
 Sample : AR1268ICC400
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 03:11:44 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ112023CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 03:09:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds						
1) SA Tetrachlo...	3.458	2.775	86161960	94202355	20.000	20.000
2) SA Decachlor...	8.689	7.570	216.5E6	247.3E6	40.000	40.000
Target Compounds						
41) L9 AR-1268-1	7.565	6.528	199.3E6	248.8E6	400.000	400.000
42) L9 AR-1268-2	7.643	6.590	179.6E6	232.9E6	400.000	400.000
43) L9 AR-1268-3	7.825	6.791	154.1E6	240.3E6	400.000	400.000
44) L9 AR-1268-4	8.151	7.084	64296929	107.5E6	400.000	400.000
45) L9 AR-1268-5	8.451	7.352	448.7E6	629.3E6	400.000	400.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112023\
Data File : PQ064225.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2023 01:15
Operator : YP\AJ
Sample : AR1268ICC400
Misc :
ALS Vial : 40 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 03:11:44 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ112023CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 03:09:16 2023
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.458 min
Delta R.T.: 0.000 min
Response: 86161960
Conc: 20.00 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 2.775 min
Delta R.T.: 0.000 min
Response: 94202355
Conc: 20.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.689 min
Delta R.T.: 0.000 min
Response: 216471490
Conc: 40.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 7.570 min
Delta R.T.: 0.000 min
Response: 247304092
Conc: 40.00 ng/ml




#41 AR-1268-1

R.T.: 7.565 min
Delta R.T.: 0.000 min
Response: 199278342
Conc: 400.00 ng/ml




#41 AR-1268-1

R.T.: 6.528 min
Delta R.T.: 0.000 min
Response: 248802703
Conc: 400.00 ng/ml




#42 AR-1268-2

R.T.: 7.643 min
Delta R.T.: 0.000 min
Response: 179595509
Conc: 400.00 ng/ml




#42 AR-1268-2

R.T.: 6.590 min
Delta R.T.: 0.000 min
Response: 232925372
Conc: 400.00 ng/ml




#43 AR-1268-3

R.T.: 7.825 min
Delta R.T.: 0.000 min
Response: 154087015
Conc: 400.00 ng/ml




#43 AR-1268-3

R.T.: 6.791 min
Delta R.T.: 0.000 min
Response: 240291637
Conc: 400.00 ng/ml




#44 AR-1268-4

R.T.: 8.151 min
Delta R.T.: 0.000 min
Response: 64296929
Conc: 400.00 ng/ml




#44 AR-1268-4

R.T.: 7.084 min
Delta R.T.: 0.000 min
Response: 107485975
Conc: 400.00 ng/ml



#45 AR-1268-5

R.T.: 8.451 min
Delta R.T.: 0.000 min
Response: 448676195
Conc: 400.00 ng/ml



#45 AR-1268-5

R.T.: 7.352 min
Delta R.T.: 0.000 min
Response: 629339522
Conc: 400.00 ng/ml