

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112023\
 Data File : PQ064217.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2023 23:17
 Operator : YP\AJ
 Sample : AR1254ICC1600
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12545008

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 01:55:16 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ112023CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 01:55:08 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.457	2.774	338.1E6	366.8E6	73.859	73.059
2) SA Decachlor...	8.689	7.570	449.8E6	556.5E6	142.389	144.884
Target Compounds						
26) L6 AR-1254-1	5.402	4.544	238.5E6	301.9E6	1386.248	1380.437
27) L6 AR-1254-2	5.621	4.692	372.6E6	258.8E6	1388.796	1364.884
28) L6 AR-1254-3	5.976	5.075	391.5E6	416.8E6	1419.207	1408.177
29) L6 AR-1254-4	6.263	5.303	280.8E6	264.9E6	1420.186	1402.052
30) L6 AR-1254-5	6.679	5.710	312.6E6	388.2E6	1430.734	1422.970

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112023\
Data File : PQ064217.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2023 23:17
Operator : YP\AJ
Sample : AR1254ICC1600
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12545008

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 01:55:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ112023CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 01:55:08 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

