

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112024\
 Data File : PQ069513.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2024 11:20
 Operator : YP\AJ
 Sample : P4847-05MS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 A0AT5MS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/21/2024
 Supervised By :Ankita Jodhani 11/21/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 01:40:43 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111824CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 20 07:44:43 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.336	2.738	159.4E6	256.1E6	21.974	23.674
2) SA Decachlor...	8.478	7.480	62080003	183.8E6	22.818	27.398
Target Compounds						
3) L1 AR-1016-1	4.427	3.729	94045826	176.7E6	421.388	485.857
4) L1 AR-1016-2	4.446	3.744	148.1E6	260.7E6	439.905	485.266
5) L1 AR-1016-3	4.502	3.903	92450305	139.9E6	447.724	484.247
6) L1 AR-1016-4	4.594	3.952	82525490	111.1E6	481.998	473.899
7) L1 AR-1016-5	4.878	4.146	69688650	143.5E6	417.242	482.193
31) L7 AR-1260-1	5.976	5.132	125.9E6	311.1E6	410.882	558.733 #
32) L7 AR-1260-2	6.236	5.322	155.8E6	328.9E6	414.184	479.057m
33) L7 AR-1260-3	6.590	5.462	84674902	309.6E6	304.861	470.400m#
34) L7 AR-1260-4	6.805	5.922	105.4E6	218.4E6	338.249m	380.482
35) L7 AR-1260-5	7.119	6.168	205.9E6	501.6E6	317.631	369.813

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112024\
Data File : PQ069513.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2024 11:20
Operator : YP\AJ
Sample : P4847-05MS
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :

ECD_Q

ClientSampleId :

A0AT5MS

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 11/21/2024
Supervised By :Ankita Jodhani 11/21/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 01:40:43 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111824CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 20 07:44:43 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

