

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112123\
 Data File : PQ064260.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2023 12:30
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 00:32:09 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 08 18:43:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.460	2.774	189.9E6	203.4E6	44.549	55.302
2) SA Decachlor...	8.693	7.571	177.3E6	261.0E6	54.907	55.711
Target Compounds						
26) L6 AR-1254-1	5.404	4.544	72921258	95267354	464.615	490.150
27) L6 AR-1254-2	5.623	4.692	108.4E6	84674390	457.212	486.063
28) L6 AR-1254-3	5.979	5.076	113.2E6	125.7E6	456.542	469.704
29) L6 AR-1254-4	6.265	5.303	82721680	81385169	456.465	463.918
30) L6 AR-1254-5	6.681	5.710	93340212	117.4E6	467.965	460.606



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112123\
Data File : PQ064260.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2023 12:30
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 22 00:32:09 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110823.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 08 18:43:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.460 min
Delta R.T.: 0.000 min
Response: 189887747
Conc: 44.55 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 2.774 min
Delta R.T.: -0.003 min
Response: 203381661
Conc: 55.30 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.693 min
Delta R.T.: -0.001 min
Response: 177275201
Conc: 54.91 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.571 min
Delta R.T.: -0.006 min
Response: 261031183
Conc: 55.71 ng/ml



#26 AR-1254-1


R.T.: 5.404 min
Delta R.T.: 0.000 min
Response: 72921258
Conc: 464.62 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 4.544 min
Delta R.T.: -0.004 min
Response: 95267354
Conc: 490.15 ng/ml




#27 AR-1254-2

R.T.: 5.623 min
Delta R.T.: 0.000 min
Response: 108430416
Conc: 457.21 ng/ml



#27 AR-1254-2


R.T.: 4.692 min
Delta R.T.: -0.004 min
Response: 84674390
Conc: 486.06 ng/ml



#28 AR-1254-3


R.T.: 5.979 min
Delta R.T.: 0.000 min
Response: 113195637
Conc: 456.54 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.076 min
Delta R.T.: -0.004 min
Response: 125725639
Conc: 469.70 ng/ml




#29 AR-1254-4

R.T.: 6.265 min
Delta R.T.: 0.000 min
Response: 82721680
Conc: 456.46 ng/ml



#29 AR-1254-4


R.T.: 5.303 min
Delta R.T.: -0.005 min
Response: 81385169
Conc: 463.92 ng/ml



#30 AR-1254-5

R.T.: 6.681 min
Delta R.T.: -0.001 min
Response: 93340212
Conc: 467.96 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 5.710 min
Delta R.T.: -0.005 min
Response: 117427238
Conc: 460.61 ng/ml