

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112219\
Data File : PQ045337.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Nov 2019 16:05
Operator : SM\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
ECD_Q
LabSampleId :
AR1254332

Manual Integrations
APPROVED

Ankita
11/25/2019 11:49:15 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 22 16:42:44 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110719CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 08 03:35:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
R.T. Response Conc
7.32 116771588 360.90
7.55 185239610 373.09
7.93 200846299 377.35
8.22 139216860 360.39
8.65 166572377 385.60

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
6.46 128110760 352.17
6.62 118185093 366.94
7.05 212024727 382.24
7.29 145178657 406.18
7.72 211990545 423.96

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112219\
Data File : PQ045337.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Nov 2019 16:05
Operator : SM\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
ECD_Q
LabSampleId :
AR1254332

Manual Integrations
APPROVED

Ankita
11/25/2019 11:49:15 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 22 16:42:44 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110719CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 08 03:35:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 #2 (L6)
R.T. Response Conc
7.32 116771588 360.90
7.55 185239610 373.09
7.93 200846299 377.35
8.22 150033224 388.39
8.65 166572377 385.60

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
6.46 128110760 352.17
6.62 118185093 366.94
7.05 212024727 382.24
7.29 145178657 406.18
7.72 211990545 423.96

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112219\
 Data File : PQ045337.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Nov 2019 16:05
 Operator : SM\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 LabSampleId :
 AR1254332

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 16:42:44 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ110719CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 08 03:35:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Manual Integrations
 APPROVED

Ankita
 11/25/2019 11:49:15 AM

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.094	4.299	116.9E6	71224845	14.732	15.105
2) SA Decachlor...	11.155	9.716	257.3E6	230.6E6	37.155	33.979
Target Compounds						
26) L6 AR-1254-1	7.316	6.461	116.8E6	128.1E6	360.900	352.165
27) L6 AR-1254-2	7.545	6.620	185.2E6	118.2E6	373.090	366.943
28) L6 AR-1254-3	7.926	7.047	200.8E6	212.0E6	377.346	382.237
29) L6 AR-1254-4	8.221	7.286	150.0E6	145.2E6	388.394m	406.181
30) L6 AR-1254-5	8.651	7.720	166.6E6	212.0E6	385.600	423.962

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

3 ^{AJ}
 11/29/19