

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Data\PQ112318\  
Data File : PQ034923.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 23 Nov 2018 18:59  
Operator : SM\SJ  
Sample : J6040-10  
Misc :  
ALS Vial : 39 Sample Multiplier: 1


Instrument :  
ECD\_Q  
ClientSampleId :  
CB4L5

Manual Integrations  
APPROVED

MOHAMMAD  
11/26/2018 2:00:56 PM

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Nov 23 22:37:45 2018  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Method\PQ111718CLP.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Fri Nov 16 03:52:57 2018  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

4.529min 18.793 ng/ml

response 51764205

(1) Tetrachloro-m-xylene #2 (SA)

3.819min 20.342 ng/ml

response 34478958

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Data\PQ112318\  
Data File : PQ034923.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 23 Nov 2018 18:59  
Operator : SM\SJ  
Sample : J6040-10  
Misc :  
ALS Vial : 39 Sample Multiplier: 1


Instrument :  
ECD\_Q  
ClientSampleId :  
CB4L5

Manual Integrations  
APPROVED

MOHAMMAD  
11/26/2018 2:00:56 PM

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Nov 23 22:37:45 2018  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Method\PQ111718CLP.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Fri Nov 16 03:52:57 2018  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

4.529min 19.739 ng/ml m

response 54368607

(1) Tetrachloro-m-xylene #2 (SA)

3.819min 21.033 ng/ml m

response 35650343

## Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Data\PQ112318\  
 Data File : PQ034923.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 23 Nov 2018 18:59  
 Operator : SM\SJ  
 Sample : J6040-10  
 Misc :  
 ALS Vial : 39 Sample Multiplier: 1

Instrument :  
 ECD\_Q  
 ClientSampleId :  
 CB4L5

Manual Integrations  
 APPROVED

MOHAMMAD  
 11/26/2018 2:00:56 PM

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Nov 23 22:37:45 2018  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Method\PQ111718CLP.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Fri Nov 16 03:52:57 2018  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2  | Resp#1   | Resp#2   | ng/ml   | ng/ml    |
|-----------------------------|--------|-------|----------|----------|---------|----------|
| -----                       |        |       |          |          |         |          |
| System Monitoring Compounds |        |       |          |          |         |          |
| 1) SA Tetrachlo...          | 4.529  | 3.819 | 54368607 | 35650343 | 19.739m | 21.033m  |
| 2) SA Decachlor...          | 10.312 | 8.837 | 60700921 | 35211455 | 22.033  | 18.066   |
| Target Compounds            |        |       |          |          |         |          |
| 31) L7 AR-1260-1            | 7.301  | 6.382 | 16273649 | 9857178  | 117.233 | 91.936   |
| 32) L7 AR-1260-2            | 7.555  | 6.569 | 55407481 | 12097616 | 324.089 | 90.080 # |
| 33) L7 AR-1260-3            | 7.916  | 6.723 | 11438045 | 11530067 | 107.325 | 93.920   |
| 34) L7 AR-1260-4            | 8.147  | 7.191 | 17353740 | 9385906  | 131.024 | 108.822  |
| 35) L7 AR-1260-5            | 8.473  | 7.432 | 27739353 | 25229020 | 102.237 | 114.771  |
| -----                       |        |       |          |          |         |          |

5511/28/18

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.