

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112420\
 Data File : PQ051221.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Nov 2020 17:44
 Operator : DD\AJ
 Sample : L4837-14
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 PCB-GPC2-BLANK-SPIKE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 25 01:28:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111820CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 19 05:59:02 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.227	4.254	14766447	-13718	0.652	N.D. #
2)	SA Decachlor...	0.000	9.613	0	2106415	N.D.	0.412 #
Target Compounds							
3)	L1 AR-1016-1	6.561	5.521	330.7E6	100.4E6	439.911	473.797
4)	L1 AR-1016-2	6.586	5.541	498.6E6	141.3E6	458.695	474.007
5)	L1 AR-1016-3	6.653	5.734	285.7E6	71639466	444.249	468.251
6)	L1 AR-1016-4	6.760	5.783	250.8E6	54269871	462.981	447.332
7)	L1 AR-1016-5	7.074	6.014	235.0E6	72895692	451.473	466.047
8)	L2 AR-1221-1	5.486	4.521	78380685	13212151	354.431	214.754 #
9)	L2 AR-1221-2	5.589	4.621	55915639	13851907	364.247	323.840
10)	L2 AR-1221-3	5.677	4.710	184.1E6	49038561	352.194	334.372
11)	L3 AR-1232-1	5.677	4.710	184.1E6	49038561	431.850	411.867
12)	L3 AR-1232-2	6.268	5.541	258.8E6	141.3E6	1151.310	1134.118
13)	L3 AR-1232-3	6.586	5.734	498.6E6	71639466	1073.674	1128.894
14)	L3 AR-1232-4	6.760	5.829	250.8E6	67568391	1081.423	1225.692
15)	L3 AR-1232-5	6.854	6.014	203.2E6	72895692	1215.942	1219.100
16)	L4 AR-1242-1	6.561	5.521	330.7E6	100.4E6	547.069	596.986
17)	L4 AR-1242-2	6.586	5.541	498.6E6	141.3E6	575.952	604.985
18)	L4 AR-1242-3	6.653	5.734	285.7E6	71639466	553.147	592.139
19)	L4 AR-1242-4	6.760	5.829	250.8E6	67568391	581.345	579.143
20)	L4 AR-1242-5	7.542	6.393	42962559	58264827	91.600	368.262 #
21)	L5 AR-1248-1	6.561	5.521	330.7E6	100.4E6	766.924	847.208
22)	L5 AR-1248-2	6.854	5.783	203.2E6	54269871	337.494	329.538
23)	L5 AR-1248-3	7.074	5.829	235.0E6	67568391	338.472	395.888
24)	L5 AR-1248-4	7.501	6.014	36523429	72895692	45.091	353.283 #
25)	L5 AR-1248-5	7.542	6.435	42962559	7673754	54.694	35.777 #
26)	L6 AR-1254-1	7.472	6.393	182.4E6	58264827	212.700	162.676
27)	L6 AR-1254-2	7.698	6.551	193.4E6	56486419	145.126	182.313 #
28)	L6 AR-1254-3	8.083	6.991	139.1E6	101.7E6	96.017	196.879 #
29)	L6 AR-1254-4	8.377	7.210	434.1E6	11378399	414.464	34.953 #
30)	L6 AR-1254-5	8.806	7.640	760.0E6	219.4E6	688.922	483.599 #
31)	L7 AR-1260-1	8.249	7.109	578.3E6	150.1E6	663.644	487.681 #
32)	L7 AR-1260-2	8.512	7.304	772.6E6	186.5E6	754.341	480.736 #
33)	L7 AR-1260-3	8.879	7.462	365.0E6	174.3E6	493.497	482.560
34)	L7 AR-1260-4	9.124	7.946	377.3E6	124.2E6	446.972	396.537
35)	L7 AR-1260-5	9.470	8.191	757.4E6	311.3E6	430.617	396.297
36)	L8 AR-1262-1	8.879	7.640	365.0E6	219.4E6	287.024	943.375 #
37)	L8 AR-1262-2	9.470	8.191	757.4E6	311.3E6	333.529	361.828
38)	L8 AR-1262-3	9.829	8.480	472.3E6	52678472	499.463	168.084 #
39)	L8 AR-1262-4	9.885	8.542	249.6E6	213.5E6	434.103	347.763
40)	L8 AR-1262-5	10.619	9.048	174.8E6	65973023	217.139	237.416
41)	L9 AR-1268-1	9.829	8.480	472.3E6	52678472	173.311	55.986 #
42)	L9 AR-1268-2	9.885	8.542	249.6E6	213.5E6	100.218	240.299 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112420\
 Data File : PQ051221.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Nov 2020 17:44
 Operator : DD\AJ
 Sample : L4837-14
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 PCB-GPC2-BLANK-SPIKE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 25 01:28:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111820CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 19 05:59:02 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43) L9	AR-1268-3	10.151	8.756	19621203	6826685	9.105	9.112
44) L9	AR-1268-4	10.619	9.048	174.8E6	65973023	187.006	210.301
45) L9	AR-1268-5	11.069	9.350	56061463	20175203	7.453	8.291



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112420\
Data File : PQ051221.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Nov 2020 17:44
Operator : DD\AJ
Sample : L4837-14
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 25 01:28:55 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111820CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 19 05:59:02 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.227 min
Delta R.T.: -0.009 min
Response: 14766447
Conc: 0.65 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#1 Tetrachloro-m-xylene

R.T.: 4.254 min
Delta R.T.: -0.002 min
Response: -13718
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 11.435 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl


R.T.: 9.613 min
Delta R.T.: -0.006 min
Response: 2106415
Conc: 0.41 ng/ml



#3 AR-1016-1


R.T.: 6.561 min
Delta R.T.: 0.002 min
Response: 330726922
Conc: 439.91 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#3 AR-1016-1

R.T.: 5.521 min
Delta R.T.: 0.000 min
Response: 100414781
Conc: 473.80 ng/ml




#4 AR-1016-2

R.T.: 6.586 min
Delta R.T.: 0.002 min
Response: 498606256
Conc: 458.69 ng/ml



#4 AR-1016-2


R.T.: 5.541 min
Delta R.T.: 0.000 min
Response: 141269894
Conc: 474.01 ng/ml



#5 AR-1016-3


R.T.: 6.653 min
Delta R.T.: 0.001 min
Response: 285681544
Conc: 444.25 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#5 AR-1016-3

R.T.: 5.734 min
Delta R.T.: 0.000 min
Response: 71639466
Conc: 468.25 ng/ml




#6 AR-1016-4

R.T.: 6.760 min
Delta R.T.: 0.000 min
Response: 250764805
Conc: 462.98 ng/ml



#6 AR-1016-4


R.T.: 5.783 min
Delta R.T.: 0.000 min
Response: 54269871
Conc: 447.33 ng/ml



#7 AR-1016-5


R.T.: 7.074 min
Delta R.T.: 0.001 min
Response: 235031377
Conc: 451.47 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#7 AR-1016-5

R.T.: 6.014 min
Delta R.T.: -0.001 min
Response: 72895692
Conc: 466.05 ng/ml




#8 AR-1221-1

R.T.: 5.486 min
Delta R.T.: 0.000 min
Response: 78380685
Conc: 354.43 ng/ml



#8 AR-1221-1


R.T.: 4.521 min
Delta R.T.: 0.006 min
Response: 13212151
Conc: 214.75 ng/ml



#9 AR-1221-2


R.T.: 5.589 min
Delta R.T.: 0.004 min
Response: 55915639
Conc: 364.25 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#9 AR-1221-2

R.T.: 4.621 min
Delta R.T.: 0.005 min
Response: 13851907
Conc: 323.84 ng/ml




#10 AR-1221-3

R.T.: 5.677 min
Delta R.T.: 0.004 min
Response: 184065983
Conc: 352.19 ng/ml



#10 AR-1221-3


R.T.: 4.710 min
Delta R.T.: 0.005 min
Response: 49038561
Conc: 334.37 ng/ml



#11 AR-1232-1


R.T.: 5.677 min
Delta R.T.: 0.003 min
Response: 184065983
Conc: 431.85 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#11 AR-1232-1

R.T.: 4.710 min
Delta R.T.: 0.005 min
Response: 49038561
Conc: 411.87 ng/ml




#12 AR-1232-2

R.T.: 6.268 min
Delta R.T.: 0.002 min
Response: 258758943
Conc: 1151.31 ng/ml



#12 AR-1232-2


R.T.: 5.541 min
Delta R.T.: 0.000 min
Response: 141269894
Conc: 1134.12 ng/ml



#13 AR-1232-3


R.T.: 6.586 min
Delta R.T.: 0.002 min
Response: 498606256
Conc: 1073.67 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#13 AR-1232-3

R.T.: 5.734 min
Delta R.T.: 0.000 min
Response: 71639466
Conc: 1128.89 ng/ml




#14 AR-1232-4

R.T.: 6.760 min
Delta R.T.: 0.000 min
Response: 250764805
Conc: 1081.42 ng/ml



#14 AR-1232-4


R.T.: 5.829 min
Delta R.T.: 0.000 min
Response: 67568391
Conc: 1225.69 ng/ml



#15 AR-1232-5


R.T.: 6.854 min
Delta R.T.: 0.000 min
Response: 203248342
Conc: 1215.94 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#15 AR-1232-5

R.T.: 6.014 min
Delta R.T.: 0.000 min
Response: 72895692
Conc: 1219.10 ng/ml




#16 AR-1242-1

R.T.: 6.561 min
Delta R.T.: 0.001 min
Response: 330726922
Conc: 547.07 ng/ml



#16 AR-1242-1


R.T.: 5.521 min
Delta R.T.: 0.000 min
Response: 100414781
Conc: 596.99 ng/ml



#17 AR-1242-2


R.T.: 6.586 min
Delta R.T.: 0.002 min
Response: 498606256
Conc: 575.95 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#17 AR-1242-2

R.T.: 5.541 min
Delta R.T.: 0.000 min
Response: 141269894
Conc: 604.98 ng/ml




#18 AR-1242-3

R.T.: 6.653 min
Delta R.T.: 0.001 min
Response: 285681544
Conc: 553.15 ng/ml



#18 AR-1242-3


R.T.: 5.734 min
Delta R.T.: 0.000 min
Response: 71639466
Conc: 592.14 ng/ml



#19 AR-1242-4


R.T.: 6.760 min
Delta R.T.: 0.000 min
Response: 250764805
Conc: 581.35 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#19 AR-1242-4

R.T.: 5.829 min
Delta R.T.: 0.000 min
Response: 67568391
Conc: 579.14 ng/ml




#20 AR-1242-5

R.T.: 7.542 min
Delta R.T.: 0.000 min
Response: 42962559
Conc: 91.60 ng/ml



#20 AR-1242-5


R.T.: 6.393 min
Delta R.T.: -0.001 min
Response: 58264827
Conc: 368.26 ng/ml



#21 AR-1248-1


R.T.: 6.561 min
Delta R.T.: 0.001 min
Response: 330726922
Conc: 766.92 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#21 AR-1248-1

R.T.: 5.521 min
Delta R.T.: 0.000 min
Response: 100414781
Conc: 847.21 ng/ml




#22 AR-1248-2

R.T.: 6.854 min
Delta R.T.: 0.000 min
Response: 203248342
Conc: 337.49 ng/ml



#22 AR-1248-2


R.T.: 5.783 min
Delta R.T.: 0.000 min
Response: 54269871
Conc: 329.54 ng/ml



#23 AR-1248-3


R.T.: 7.074 min
Delta R.T.: 0.000 min
Response: 235031377
Conc: 338.47 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#23 AR-1248-3

R.T.: 5.829 min
Delta R.T.: 0.000 min
Response: 67568391
Conc: 395.89 ng/ml




#24 AR-1248-4

R.T.: 7.501 min
Delta R.T.: 0.000 min
Response: 36523429
Conc: 45.09 ng/ml



#24 AR-1248-4


R.T.: 6.014 min
Delta R.T.: -0.002 min
Response: 72895692
Conc: 353.28 ng/ml



#25 AR-1248-5


R.T.: 7.542 min
Delta R.T.: 0.001 min
Response: 42962559
Conc: 54.69 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#25 AR-1248-5

R.T.: 6.435 min
Delta R.T.: -0.003 min
Response: 7673754
Conc: 35.78 ng/ml




#26 AR-1254-1

R.T.: 7.472 min
Delta R.T.: 0.001 min
Response: 182409087
Conc: 212.70 ng/ml



#26 AR-1254-1


R.T.: 6.393 min
Delta R.T.: 0.000 min
Response: 58264827
Conc: 162.68 ng/ml



#27 AR-1254-2


R.T.: 7.698 min
Delta R.T.: 0.000 min
Response: 193422154
Conc: 145.13 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#27 AR-1254-2

R.T.: 6.551 min
Delta R.T.: -0.001 min
Response: 56486419
Conc: 182.31 ng/ml




#28 AR-1254-3

R.T.: 8.083 min
Delta R.T.: 0.001 min
Response: 139050112
Conc: 96.02 ng/ml



#28 AR-1254-3


R.T.: 6.991 min
Delta R.T.: 0.017 min
Response: 101736351
Conc: 196.88 ng/ml



#29 AR-1254-4


R.T.: 8.377 min
Delta R.T.: 0.001 min
Response: 434114792
Conc: 414.46 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#29 AR-1254-4

R.T.: 7.210 min
Delta R.T.: -0.002 min
Response: 11378399
Conc: 34.95 ng/ml




#30 AR-1254-5

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 759965731
Conc: 688.92 ng/ml



#30 AR-1254-5


R.T.: 7.640 min
Delta R.T.: -0.002 min
Response: 219438831
Conc: 483.60 ng/ml



#31 AR-1260-1


R.T.: 8.249 min
Delta R.T.: 0.000 min
Response: 578274608
Conc: 663.64 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#31 AR-1260-1

R.T.: 7.109 min
Delta R.T.: -0.003 min
Response: 150089201
Conc: 487.68 ng/ml




#32 AR-1260-2

R.T.: 8.512 min
Delta R.T.: 0.000 min
Response: 772646273
Conc: 754.34 ng/ml



#32 AR-1260-2


R.T.: 7.304 min
Delta R.T.: -0.003 min
Response: 186534714
Conc: 480.74 ng/ml



#33 AR-1260-3


R.T.: 8.879 min
Delta R.T.: 0.000 min
Response: 365001523
Conc: 493.50 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#33 AR-1260-3

R.T.: 7.462 min
Delta R.T.: -0.004 min
Response: 174326056
Conc: 482.56 ng/ml




#34 AR-1260-4

R.T.: 9.124 min
Delta R.T.: 0.000 min
Response: 377328878
Conc: 446.97 ng/ml



#34 AR-1260-4


R.T.: 7.946 min
Delta R.T.: -0.004 min
Response: 124209691
Conc: 396.54 ng/ml



#35 AR-1260-5


R.T.: 9.470 min
Delta R.T.: 0.000 min
Response: 757350949
Conc: 430.62 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#35 AR-1260-5

R.T.: 8.191 min
Delta R.T.: -0.004 min
Response: 311332247
Conc: 396.30 ng/ml




#36 AR-1262-1

R.T.: 8.879 min
Delta R.T.: 0.002 min
Response: 365001523
Conc: 287.02 ng/ml



#36 AR-1262-1


R.T.: 7.640 min
Delta R.T.: -0.001 min
Response: 219438831
Conc: 943.38 ng/ml



#37 AR-1262-2


R.T.: 9.470 min
Delta R.T.: 0.003 min
Response: 757350949
Conc: 333.53 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#37 AR-1262-2

R.T.: 8.191 min
Delta R.T.: -0.002 min
Response: 311332247
Conc: 361.83 ng/ml




#38 AR-1262-3

R.T.: 9.829 min
Delta R.T.: 0.028 min
Response: 472317026
Conc: 499.46 ng/ml



#38 AR-1262-3


R.T.: 8.480 min
Delta R.T.: -0.002 min
Response: 52678472
Conc: 168.08 ng/ml



#39 AR-1262-4


R.T.: 9.885 min
Delta R.T.: -0.017 min
Response: 249638850
Conc: 434.10 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#39 AR-1262-4

R.T.: 8.542 min
Delta R.T.: -0.003 min
Response: 213457240
Conc: 347.76 ng/ml




#40 AR-1262-5

R.T.: 10.619 min
Delta R.T.: 0.000 min
Response: 174789339
Conc: 217.14 ng/ml



#40 AR-1262-5


R.T.: 9.048 min
Delta R.T.: 0.000 min
Response: 65973023
Conc: 237.42 ng/ml



#41 AR-1268-1


R.T.: 9.829 min
Delta R.T.: 0.029 min
Response: 472317026
Conc: 173.31 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#41 AR-1268-1

R.T.: 8.480 min
Delta R.T.: 0.000 min
Response: 52678472
Conc: 55.99 ng/ml




#42 AR-1268-2

R.T.: 9.885 min
Delta R.T.: -0.017 min
Response: 249638850
Conc: 100.22 ng/ml



#42 AR-1268-2


R.T.: 8.542 min
Delta R.T.: -0.003 min
Response: 213457240
Conc: 240.30 ng/ml



#43 AR-1268-3


R.T.: 10.151 min
Delta R.T.: 0.005 min
Response: 19621203
Conc: 9.10 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE




#43 AR-1268-3

R.T.: 8.756 min
Delta R.T.: 0.000 min
Response: 6826685
Conc: 9.11 ng/ml




#44 AR-1268-4

R.T.: 10.619 min
Delta R.T.: 0.003 min
Response: 174789339
Conc: 187.01 ng/ml



#44 AR-1268-4


R.T.: 9.048 min
Delta R.T.: 0.000 min
Response: 65973023
Conc: 210.30 ng/ml



#45 AR-1268-5

R.T.: 11.069 min
Delta R.T.: 0.003 min
Response: 56061463
Conc: 7.45 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-GPC2-BLANK-SPIKE



#45 AR-1268-5

R.T.: 9.350 min
Delta R.T.: 0.000 min
Response: 20175203
Conc: 8.29 ng/ml