

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112420\
 Data File : PQ051252.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Nov 2020 02:45
 Operator : DD\AJ
 Sample : L4855-10
 Misc :
 ALS Vial : 81 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 CB7Z6

Manual Integrations
 APPROVED

Ankita
 11/25/2020 5:38:30 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 25 06:26:32 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111820CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 19 05:59:02 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.236	4.254	490.4E6	132.7E6	21.658	22.258
2) SA Decachlor...	11.430	9.609	997.8E6	330.0E6	60.809	64.629
Target Compounds						
31) L7 AR-1260-1	8.246	7.107	140.1E6	57547950	160.751	186.989
32) L7 AR-1260-2	8.509	7.302	228.4E6	55091396	222.959	141.981m#
33) L7 AR-1260-3	8.876	7.460	149.4E6	46298440	201.967	128.161m#
34) L7 AR-1260-4	9.120	7.943	165.2E6	50289104	195.706	160.547
35) L7 AR-1260-5	9.467	8.187	342.2E6	121.0E6	194.554	153.981

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112420\
Data File : PQ051252.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Nov 2020 02:45
Operator : DD\AJ
Sample : L4855-10
Misc :
ALS Vial : 81 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
CB7Z6

Manual Integrations
APPROVED

Ankita
11/25/2020 5:38:30 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 25 06:26:32 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ11820CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 19 05:59:02 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

