Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112520\
Data File : PQ051326.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Nov 2020 23:34
Operator : DD\AJ
Sample : L4866-01
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
C0B74

Manual Integrations APPROVED
Ankita
11/27/2020 12:53:34 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 26 08:07:14 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111820CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 19 05:59:02 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	5.235	4.254	284.2E6	75200673	12.550	12.613
2)	SA Decachlor...	11.430	9.606	322.2E6	102.9E6	19.635	20.154m
	Target Compounds						
26)	L6 AR-1254-1	7.467	6.389	43412518	13337216	50.622m	37.238m#
27)	L6 AR-1254-2	7.696	6.547	62699675	12468917	47.044m	40.244m
28)	L6 AR-1254-3	8.081	6.970	16911411	19346258	11.678m	37.439m#
29)	L6 AR-1254-4	8.374	7.205	60544548	13314301	57.804m	40.900m#
30)	L6 AR-1254-5	8.804	7.637	84800282	33264015	76.873m	73.307m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112520\
Data File : PQ051326.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Nov 2020 23:34
Operator : DD\AJ
Sample : L4866-01
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
C0B74

Manual Integrations APPROVED
Ankita
11/27/2020 12:53:34 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 26 08:07:14 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111820CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 19 05:59:02 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm