

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\pq112718\
 Data File : PQ035008.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Nov 2018 07:15
 Operator : SM\SJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 28 00:39:12 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111718CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 16 03:52:57 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds

7) L1	AR-1016-5	0.000	5.330	0	966474	N.D.	18.645 #
15) L3	AR-1232-5	0.000	5.330	0	966474	N.D.	49.932 #
20) L4	AR-1242-5	6.581	0.000	4161877	0	67.259	N.D. #
24) L5	AR-1248-4	6.581	5.330	4161877	966474	37.945	13.927 #
25) L5	AR-1248-5	6.581	0.000	4161877	0	39.926	N.D. #
26) L6	AR-1254-1	6.581	0.000	4161877	0	37.201	N.D. #
29) L6	AR-1254-4	7.403	0.000	2635462	0	18.567	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\pq112718\
Data File : PQ035008.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Nov 2018 07:15
Operator : SM\SJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 28 00:39:12 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ111718CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 16 03:52:57 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#7 AR-1016-5


R.T.: 0.000 min
Exp R.T. : 6.191 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :




#7 AR-1016-5

R.T.: 5.330 min
Delta R.T.: -0.030 min
Response: 966474
Conc: 18.65 ng/ml




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 5.985 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 5.330 min
Delta R.T.: -0.029 min
Response: 966474
Conc: 49.93 ng/ml



#20 AR-1242-5


R.T.: 6.581 min
Delta R.T.: -0.052 min
Response: 4161877
Conc: 67.26 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T. : 5.709 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 6.581 min
Delta R.T.: -0.013 min
Response: 4161877
Conc: 37.94 ng/ml



#24 AR-1248-4


R.T.: 5.330 min
Delta R.T.: -0.030 min
Response: 966474
Conc: 13.93 ng/ml



#25 AR-1248-5


R.T.: 6.581 min
Delta R.T.: -0.053 min
Response: 4161877
Conc: 39.93 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T. : 5.751 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 6.581 min
Delta R.T.: 0.014 min
Response: 4161877
Conc: 37.20 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T. : 5.710 min
Response: 0
Conc: N.D.



#29 AR-1254-4

R.T.: 7.403 min
Delta R.T.: -0.035 min
Response: 2635462
Conc: 18.57 ng/ml

Instrument :
ECD_Q
ClientSampleId :



#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.489 min
Response: 0
Conc: N.D.