

## Exhibit D – Section 17

TABLE 2. CONCENTRATION LEVELS OF INITIAL CALIBRATION AND CONTINUING CALIBRATION VERIFICATION AND TECHNICAL ACCEPTANCE CRITERIA FOR PESTICIDES

| Analyte                          | Concentration (ng/mL) |      |      |      |      | Maximum %RSD | Opening Maximum %D | Closing Maximum %D |
|----------------------------------|-----------------------|------|------|------|------|--------------|--------------------|--------------------|
|                                  | CS1                   | CS2  | CS3  | CS4  | CS5  |              |                    |                    |
| alpha-BHC                        | 5.0                   | 10.  | 20.  | 40.  | 80.  | 25.0         | ±25.0              | ±25.0              |
| gamma-BHC                        | 5.0                   | 10.  | 20.  | 40.  | 80.  | 20.0         | ±25.0              | ±25.0              |
| Heptachlor                       | 5.0                   | 10.  | 20.  | 40.  | 80.  | 20.0         | ±25.0              | ±25.0              |
| Endosulfan I                     | 5.0                   | 10.  | 20.  | 40.  | 80.  | 20.0         | ±25.0              | ±25.0              |
| Dieldrin                         | 10.                   | 20.  | 40.  | 80.  | 160  | 20.0         | ±25.0              | ±25.0              |
| Endrin                           | 10.                   | 20.  | 40.  | 80.  | 160  | 20.0         | ±25.0              | ±25.0              |
| 4,4'-DDD                         | 10.                   | 20.  | 40.  | 80.  | 160  | 20.0         | ±25.0              | ±25.0              |
| 4,4'-DDT                         | 10.                   | 20.  | 40.  | 80.  | 160  | 20.0         | ±25.0              | ±25.0              |
| Methoxychlor                     | 50.                   | 100  | 200  | 400  | 800  | 20.0         | ±25.0              | ±25.0              |
| beta-BHC                         | 5.0                   | 10.  | 20.  | 40.  | 80.  | 20.0         | ±25.0              | ±25.0              |
| delta-BHC                        | 5.0                   | 10.  | 20.  | 40.  | 80.  | 25.0         | ±25.0              | ±25.0              |
| Aldrin                           | 5.0                   | 10.  | 20.  | 40.  | 80.  | 20.0         | ±25.0              | ±25.0              |
| Heptachlor-epoxide               | 5.0                   | 10.  | 20.  | 40.  | 80.  | 20.0         | ±25.0              | ±25.0              |
| 4,4'-DDE                         | 10.                   | 20.  | 40.  | 80.  | 160  | 20.0         | ±25.0              | ±25.0              |
| Endosulfan II                    | 10.                   | 20.  | 40.  | 80.  | 160  | 20.0         | ±25.0              | ±25.0              |
| Endosulfan sulfate               | 10.                   | 20.  | 40.  | 80.  | 160  | 20.0         | ±25.0              | ±25.0              |
| Endrin ketone                    | 10.                   | 20.  | 40.  | 80.  | 160  | 20.0         | ±25.0              | ±25.0              |
| Endrin aldehyde                  | 10.                   | 20.  | 40.  | 80.  | 160  | 20.0         | ±25.0              | ±25.0              |
| cis-Chlordane                    | 5.0                   | 10.  | 20.  | 40.  | 80.  | 20.0         | ±25.0              | ±25.0              |
| trans-Chlordane                  | 5.0                   | 10.  | 20.  | 40.  | 80.  | 20.0         | ±25.0              | ±25.0              |
| Toxaphene                        | 500                   | 1000 | 2000 | 4000 | 8000 | 30.0         | ±25.0              | ±25.0              |
| Tetrachloro-m-xylene (surrogate) | 5.0                   | 10.  | 20.  | 40.  | 80.  | 20.0         | ±30.0              | ±30.0              |
| Decachlorobiphenyl (surrogate)   | 10.                   | 20.  | 40.  | 80.  | 160  | 20.0         | ±30.0              | ±30.0              |

NOTE: Only the exo-epoxy isomer (Isomer B) of heptachlor epoxide is used as an analytical standard.