

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ113018\
 Data File : PQ035144.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Nov 2018 21:34
 Operator : SM\SJ
 Sample : J6038-13 5X (Sig #1); J6038-13 (Sig #2)
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 CB4J7

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 01 00:38:55 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ113018CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 30 22:48:24 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.455	3.755	9196232	6686857	2.263	3.222 #
2) SA Decachlor...	10.186	8.750	17485809	11012229	6.891	7.108
Target Compounds						
31) L7 AR-1260-1	7.222	6.312	1066.0E6	628.2E6	7088.365	5539.979
32) L7 AR-1260-2	7.478	6.499	1203.7E6	800.5E6	6884.794	5748.528
33) L7 AR-1260-3	7.837	6.653	844.7E6	712.0E6	8020.464	5585.199 #
34) L7 AR-1260-4	8.066	7.121	947.9E6	570.8E6	7250.898	6872.268
35) L7 AR-1260-5	8.387	7.363	1896.9E6	1367.0E6	7176.383	6756.192

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ113018\
Data File : PQ035144.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Nov 2018 21:34
Operator : SM\SJ
Sample : J6038-13 5X (Sig #1); J6038-13 (Sig #2)
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
CB4J7

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 01 00:38:55 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ113018CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 30 22:48:24 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

