

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ113021\
 Data File : PQ055239.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Dec 2021 01:16
 Operator : AJ\MA
 Sample : M4068-08
 Misc : AR1248 LOQ 50 PPB
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOQ-WATER-02-QT4-2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 01 02:28:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ112921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 30 02:09:05 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.246	4.320	336.5E6	197.0E6	19.140	19.160
2) SA Decachlor...	11.435	9.724	345.0E6	181.5E6	21.871	20.266
Target Compounds						
21) L5 AR-1248-1	6.569	5.591	17831060	9880267	52.429	49.364
22) L5 AR-1248-2	6.863	5.856	25618725	15051694	55.899	53.533
23) L5 AR-1248-3	7.082	5.902	28591056	15220983	53.015	52.026
24) L5 AR-1248-4	7.511	6.089	33554098	16481201	55.297	47.346
25) L5 AR-1248-5	7.551	6.514	33229399	17733525	54.295	50.963

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ113021\
Data File : PQ055239.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Dec 2021 01:16
Operator : AJ\MA
Sample : M4068-08
Misc : AR1248 LOQ 50 PPB
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT4-2021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 01 02:28:50 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ112921.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 30 02:09:05 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

