

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ120218.M
 Title : GC EXTRACTABLES
 Last Update : Mon Dec 03 08:50:54 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ035176.D 1000 =PQ035174.D 750 =PQ035175.D
 250 =PQ035177.D 50 =PQ035178.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.338	3.207	3.541	3.588	3.416	3.418	E6	4.51
2) SA	Decachlorobiphenyl	3.717	3.715	3.710	3.552	3.025	3.544	E6	8.42
3) L1	AR-1016-1	1.122	1.074	1.147	1.240	1.400	1.197	E5	10.77
4) L1	AR-1016-2	1.734	1.651	1.794	1.924	2.052	1.831	E5	8.67
5) L1	AR-1016-3	1.070	0.992	1.072	1.216	1.257	1.122	E5	9.88
6) L1	AR-1016-4	8.651	8.230	8.905	9.506	9.903	9.039	E4	7.40
7) L1	AR-1016-5	0.870	0.830	0.877	1.006	1.028	0.922	E5	9.63
8) L2	AR-1221-1	3.400					3.400	E4	0.00
9) L2	AR-1221-2	2.521					2.521	E4	0.00
10) L2	AR-1221-3	8.265					8.265	E4	0.00
11) L3	AR-1232-1	6.697					6.697	E4	0.00
12) L3	AR-1232-2	3.557					3.557	E4	0.00
13) L3	AR-1232-3	7.984					7.984	E4	0.00
14) L3	AR-1232-4	3.958					3.958	E4	0.00
15) L3	AR-1232-5	2.935					2.935	E4	0.00
16) L4	AR-1242-1	1.009					1.009	E5	0.00
17) L4	AR-1242-2	1.520					1.520	E5	0.00
18) L4	AR-1242-3	9.053					9.053	E4	0.00
19) L4	AR-1242-4	7.445					7.445	E4	0.00
20) L4	AR-1242-5	8.489					8.489	E4	0.00
21) L5	AR-1248-1	7.482					7.482	E4	0.00
22) L5	AR-1248-2	1.072					1.072	E5	0.00
23) L5	AR-1248-3	1.258					1.258	E5	0.00
24) L5	AR-1248-4	1.437					1.437	E5	0.00
25) L5	AR-1248-5	1.369					1.369	E5	0.00
26) L6	AR-1254-1	1.575					1.575	E5	0.00
27) L6	AR-1254-2	2.523					2.523	E5	0.00
28) L6	AR-1254-3	2.710					2.710	E5	0.00
29) L6	AR-1254-4	1.898					1.898	E5	0.00
30) L6	AR-1254-5	1.923					1.923	E5	0.00
31) L7	AR-1260-1	1.863	1.768	1.860	2.075	2.424	1.998	E5	13.18
32) L7	AR-1260-2	2.236	2.134	2.237	2.426	2.422	2.291	E5	5.61
33) L7	AR-1260-3	1.380	1.326	1.378	1.483	1.458	1.405	E5	4.58
34) L7	AR-1260-4	1.763	1.728	1.763	1.842	1.901	1.799	E5	3.93
35) L7	AR-1260-5	3.561	3.536	3.602	3.658	3.412	3.554	E5	2.58
36) L8	AR-1262-1	2.156					2.156	E5	0.00
37) L8	AR-1262-2	3.666					3.666	E5	0.00
38) L8	AR-1262-3	2.291					2.291	E5	0.00
39) L8	AR-1262-4	1.712					1.712	E5	0.00
40) L8	AR-1262-5	1.118					1.118	E5	0.00
41) L9	AR-1268-1	5.545					5.545	E5	0.00
42) L9	AR-1268-2	5.059					5.059	E5	0.00
43) L9	AR-1268-3	4.280					4.280	E5	0.00
44) L9	AR-1268-4	1.591					1.591	E5	0.00
45) L9	AR-1268-5	1.282					1.282	E6	0.00

Signal #2 Calibration Files

500 =PQ035176.D 1000 =PQ035174.D 750 =PQ035175.D
 250 =PQ035177.D 50 =PQ035178.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.837	1.774	1.970	2.033	1.964	1.916	E6	5.54
2) SA	Decachlorobiphenyl	2.192	2.163	2.182	2.121	1.885	2.109	E6	6.08
3) L1	AR-1016-1	6.665	6.284	6.765	7.381	8.390	7.097	E4	11.60
4) L1	AR-1016-2	1.000	0.946	1.032	1.115	1.073	1.033	E5	6.33
5) L1	AR-1016-3	5.301	4.863	5.306	5.706	5.661	5.368	E4	6.35

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ120218.M
 Title : GC EXTRACTABLES
 Last Update : Mon Dec 03 08:50:54 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ035176.D 1000 =PQ035174.D 750 =PQ035175.D
 250 =PQ035177.D 50 =PQ035178.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	4.217	3.843	4.173	4.678	4.649	4.312	E4	8.16
7)	L1 AR-1016-5	5.557	5.172	5.559	6.263	6.224	5.755	E4	8.23
8)	L2 AR-1221-1	1.910					1.910	E4	0.00
9)	L2 AR-1221-2	1.432					1.432	E4	0.00
10)	L2 AR-1221-3	4.913					4.913	E4	0.00
11)	L3 AR-1232-1	4.026					4.026	E4	0.00
12)	L3 AR-1232-2	4.797					4.797	E4	0.00
13)	L3 AR-1232-3	2.397					2.397	E4	0.00
14)	L3 AR-1232-4	2.119					2.119	E4	0.00
15)	L3 AR-1232-5	2.352					2.352	E4	0.00
16)	L4 AR-1242-1	6.146					6.146	E4	0.00
17)	L4 AR-1242-2	9.114					9.114	E4	0.00
18)	L4 AR-1242-3	4.780					4.780	E4	0.00
19)	L4 AR-1242-4	4.606					4.606	E4	0.00
20)	L4 AR-1242-5	5.884					5.884	E4	0.00
21)	L5 AR-1248-1	4.615					4.615	E4	0.00
22)	L5 AR-1248-2	6.236					6.236	E4	0.00
23)	L5 AR-1248-3	6.389					6.389	E4	0.00
24)	L5 AR-1248-4	8.061					8.061	E4	0.00
25)	L5 AR-1248-5	8.247					8.247	E4	0.00
26)	L6 AR-1254-1	1.410					1.410	E5	0.00
27)	L6 AR-1254-2	1.184					1.184	E5	0.00
28)	L6 AR-1254-3	2.047					2.047	E5	0.00
29)	L6 AR-1254-4	1.280					1.280	E5	0.00
30)	L6 AR-1254-5	1.689					1.689	E5	0.00
31)	L7 AR-1260-1	1.267	1.180	1.266	1.366	1.400	1.296	E5	6.78
32)	L7 AR-1260-2	1.573	1.478	1.556	1.710	1.781	1.620	E5	7.61
33)	L7 AR-1260-3	1.477	1.399	1.477	1.617	1.575	1.509	E5	5.75
34)	L7 AR-1260-4	1.063	1.002	1.045	1.141	1.218	1.094	E5	7.84
35)	L7 AR-1260-5	2.780	2.677	2.774	2.832	2.511	2.715	E5	4.68
36)	L8 AR-1262-1	1.609					1.609	E5	0.00
37)	L8 AR-1262-2	2.796					2.796	E5	0.00
38)	L8 AR-1262-3	9.987					9.987	E4	0.00
39)	L8 AR-1262-4	1.901					1.901	E5	0.00
40)	L8 AR-1262-5	7.658					7.658	E4	0.00
41)	L9 AR-1268-1	4.010					4.010	E5	0.00
42)	L9 AR-1268-2	3.588					3.588	E5	0.00
43)	L9 AR-1268-3	2.865					2.865	E5	0.00
44)	L9 AR-1268-4	1.100					1.100	E5	0.00
45)	L9 AR-1268-5	8.033					8.033	E5	0.00

(#) = Out of Range