

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ120218\
 Data File : PQ035184.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Dec 2018 17:13
 Operator : SM\SJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 08:42:59 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ120218.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Dec 03 08:42:49 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.450	3.752	166.9E6	95917705	48.829	49.653
2)	SA Decachlor...	10.178	8.747	115.2E6	71503590	35.335	36.548
Target Compounds							
36)	L8 AR-1262-1	7.833	6.859	107.8E6	80468047	500.000	500.000
37)	L8 AR-1262-2	8.382	7.360	183.3E6	139.8E6	500.000	500.000
38)	L8 AR-1262-3	8.690	7.640	114.6E6	49936659	500.000	500.000
39)	L8 AR-1262-4	8.783	7.705	85603376	95066561	500.000	500.000
40)	L8 AR-1262-5	9.440	8.203	55887180	38288155	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ120218\
Data File : PQ035184.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Dec 2018 17:13
Operator : SM\SJ
Sample : AR1262ICC500
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 08:42:59 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ120218.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Dec 03 08:42:49 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

