

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ120319\
 Data File : PQ045571.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Dec 2019 21:30
 Operator : SM\AJ
 Sample : AR1242ICC100
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1242101

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 04 03:44:58 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ120319CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 04 03:44:32 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.238	4.413	36887841	32660455	5.699	5.862
2)	SA Decachlor...	11.415	9.887	126.2E6	68227233	9.388	10.546
Target Compounds							
16)	L4 AR-1242-1	6.556	5.701	24325330	15608178	125.937	122.368
17)	L4 AR-1242-2	6.580	5.722	36132319	24633311	124.590	112.328
18)	L4 AR-1242-3	6.647	5.920	21065767	12474810	123.568	120.873
19)	L4 AR-1242-4	6.753	6.013	17057423	13238560	119.727	121.008
20)	L4 AR-1242-5	7.536	6.585	21257873	15985829	123.130	119.650

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ120319\
Data File : PQ045571.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Dec 2019 21:30
Operator : SM\AJ
Sample : AR1242ICC100
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242101

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 04 03:44:58 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ120319CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 04 03:44:32 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

