

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ120319\
 Data File : PQ045579.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Dec 2019 23:40
 Operator : SM\AJ
 Sample : AR1248ICC800
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1248401

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 04 04:01:13 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ120319CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 04 03:59:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.239	4.414	278.6E6	238.7E6	38.264	38.329
2)	SA Decachlor...	11.410	9.881	1091.7E6	494.7E6	76.574	73.392
Target Compounds							
21)	L5 AR-1248-1	6.555	5.700	120.8E6	81682923	744.954	746.965
22)	L5 AR-1248-2	6.849	5.966	168.6E6	124.4E6	735.612	748.702
23)	L5 AR-1248-3	7.068	6.013	194.8E6	129.7E6	738.049	747.345
24)	L5 AR-1248-4	7.495	6.203	238.3E6	154.1E6	739.476	757.899
25)	L5 AR-1248-5	7.536	6.629	224.1E6	144.7E6	739.240	762.652

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ120319\
Data File : PQ045579.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Dec 2019 23:40
Operator : SM\AJ
Sample : AR1248ICC800
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248401

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 04 04:01:13 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ120319CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 04 03:59:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

