

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ120518.M
 Title : GC EXTRACTABLES
 Last Update : Tue Dec 04 11:50:47 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ035219.D 1000 =PQ035217.D 750 =PQ035218.D
 250 =PQ035220.D 50 =PQ035221.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	4.110	3.691	3.783	4.619	3.787	3.998 E6	9.55
2) SA	Decachlorobiphenyl	3.790	2.854	3.624	4.288	4.408	3.793 E6	16.33
3) L1	AR-1016-1	1.396	1.198	1.253	1.581	1.611	1.408 E5	13.26
4) L1	AR-1016-2	2.125	1.752	1.940	2.467	2.397	2.136 E5	14.11
5) L1	AR-1016-3	1.253	1.090	1.140	1.510	1.455	1.290 E5	14.47
6) L1	AR-1016-4	1.040	0.900	0.941	1.194	1.208	1.057 E5	13.36
7) L1	AR-1016-5	1.046	0.881	0.936	1.186	1.206	1.051 E5	13.84
8) L2	AR-1221-1	3.203					3.203 E4	0.00
9) L2	AR-1221-2	2.508					2.508 E4	0.00
10) L2	AR-1221-3	7.748					7.748 E4	0.00
11) L3	AR-1232-1	6.993					6.993 E4	0.00
12) L3	AR-1232-2	3.870					3.870 E4	0.00
13) L3	AR-1232-3	8.105					8.105 E4	0.00
14) L3	AR-1232-4	3.885					3.885 E4	0.00
15) L3	AR-1232-5	2.921					2.921 E4	0.00
16) L4	AR-1242-1	1.015					1.015 E5	0.00
17) L4	AR-1242-2	1.566					1.566 E5	0.00
18) L4	AR-1242-3	9.249					9.249 E4	0.00
19) L4	AR-1242-4	7.724					7.724 E4	0.00
20) L4	AR-1242-5	8.799					8.799 E4	0.00
21) L5	AR-1248-1	7.258					7.258 E4	0.00
22) L5	AR-1248-2	1.080					1.080 E5	0.00
23) L5	AR-1248-3	1.264					1.264 E5	0.00
24) L5	AR-1248-4	1.477					1.477 E5	0.00
25) L5	AR-1248-5	1.404					1.404 E5	0.00
26) L6	AR-1254-1	1.507					1.507 E5	0.00
27) L6	AR-1254-2	2.380					2.380 E5	0.00
28) L6	AR-1254-3	2.671					2.671 E5	0.00
29) L6	AR-1254-4	1.950					1.950 E5	0.00
30) L6	AR-1254-5	2.150					2.150 E5	0.00
31) L7	AR-1260-1	2.294	1.906	2.073	2.598	2.501	2.274 E5	12.68
32) L7	AR-1260-2	2.784	2.287	2.531	3.134	2.970	2.741 E5	12.35
33) L7	AR-1260-3	1.683	1.358	1.513	1.869	1.843	1.653 E5	13.21
34) L7	AR-1260-4	2.090	1.711	1.863	2.412	2.461	2.108 E5	15.64
35) L7	AR-1260-5	4.117	3.337	3.843	4.591	4.428	4.063 E5	12.24
36) L8	AR-1262-1	2.509					2.509 E5	0.00
37) L8	AR-1262-2	4.665					4.665 E5	0.00
38) L8	AR-1262-3	2.969					2.969 E5	0.00
39) L8	AR-1262-4	1.254					1.254 E5	0.00
40) L8	AR-1262-5	1.499					1.499 E5	0.00
41) L9	AR-1268-1	6.082					6.082 E5	0.00
42) L9	AR-1268-2	5.544					5.544 E5	0.00
43) L9	AR-1268-3	4.717					4.717 E5	0.00
44) L9	AR-1268-4	1.851					1.851 E5	0.00
45) L9	AR-1268-5	1.545					1.545 E6	0.00

Signal #2 Calibration Files

500 =PQ035219.D 1000 =PQ035217.D 750 =PQ035218.D
 250 =PQ035220.D 50 =PQ035221.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.309	2.052	2.108	2.568	1.955	2.198 E6	11.08
2) SA	Decachlorobiphenyl	2.327	1.740	2.196	2.645	2.685	2.318 E6	16.58
3) L1	AR-1016-1	8.446	7.200	7.558	9.554	9.006	8.353 E4	11.73
4) L1	AR-1016-2	1.271	1.088	1.141	1.417	1.388	1.261 E5	11.55
5) L1	AR-1016-3	6.537	5.583	5.929	7.398	7.376	6.565 E4	12.57

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ120518.M
 Title : GC EXTRACTABLES
 Last Update : Tue Dec 04 11:50:47 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ035219.D 1000 =PQ035217.D 750 =PQ035218.D
 250 =PQ035220.D 50 =PQ035221.D

	Compound	500	1000	750	250	50	Avg	%RSD
6)	L1 AR-1016-4	5.269	4.356	4.645	6.077	5.664	5.202 E4	13.63
7)	L1 AR-1016-5	7.049	5.929	6.226	8.019	7.234	6.892 E4	12.10
8)	L2 AR-1221-1	1.820					1.820 E4	0.00
9)	L2 AR-1221-2	1.329					1.329 E4	0.00
10)	L2 AR-1221-3	4.542					4.542 E4	0.00
11)	L3 AR-1232-1	4.036					4.036 E4	0.00
12)	L3 AR-1232-2	4.936					4.936 E4	0.00
13)	L3 AR-1232-3	2.501					2.501 E4	0.00
14)	L3 AR-1232-4	2.219					2.219 E4	0.00
15)	L3 AR-1232-5	2.441					2.441 E4	0.00
16)	L4 AR-1242-1	6.222					6.222 E4	0.00
17)	L4 AR-1242-2	9.369					9.369 E4	0.00
18)	L4 AR-1242-3	4.926					4.926 E4	0.00
19)	L4 AR-1242-4	4.790					4.790 E4	0.00
20)	L4 AR-1242-5	6.708					6.708 E4	0.00
21)	L5 AR-1248-1	4.588					4.588 E4	0.00
22)	L5 AR-1248-2	6.388					6.388 E4	0.00
23)	L5 AR-1248-3	6.571					6.571 E4	0.00
24)	L5 AR-1248-4	8.327					8.327 E4	0.00
25)	L5 AR-1248-5	8.661					8.661 E4	0.00
26)	L6 AR-1254-1	1.354					1.354 E5	0.00
27)	L6 AR-1254-2	1.185					1.185 E5	0.00
28)	L6 AR-1254-3	2.054					2.054 E5	0.00
29)	L6 AR-1254-4	1.320					1.320 E5	0.00
30)	L6 AR-1254-5	1.829					1.829 E5	0.00
31)	L7 AR-1260-1	1.531	1.259	1.365	1.761	1.630	1.509 E5	13.33
32)	L7 AR-1260-2	1.903	1.559	1.714	2.180	2.131	1.897 E5	14.02
33)	L7 AR-1260-3	1.783	1.483	1.621	2.015	1.855	1.751 E5	11.78
34)	L7 AR-1260-4	1.213	0.971	1.101	1.375	1.307	1.194 E5	13.54
35)	L7 AR-1260-5	3.128	2.494	2.866	3.439	3.119	3.009 E5	11.71
36)	L8 AR-1262-1	1.813					1.813 E5	0.00
37)	L8 AR-1262-2	3.374					3.374 E5	0.00
38)	L8 AR-1262-3	1.246					1.246 E5	0.00
39)	L8 AR-1262-4	2.405					2.405 E5	0.00
40)	L8 AR-1262-5	1.044					1.044 E5	0.00
41)	L9 AR-1268-1	4.325					4.325 E5	0.00
42)	L9 AR-1268-2	3.908					3.908 E5	0.00
43)	L9 AR-1268-3	3.188					3.188 E5	0.00
44)	L9 AR-1268-4	1.293					1.293 E5	0.00
45)	L9 AR-1268-5	9.908					9.908 E5	0.00

(#) = Out of Range