

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ120519\
 Data File : PQ045657.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2019 01:27
 Operator : SM\AJ
 Sample : AR1254ICC100
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254101

Manual Integrations
 APPROVED

mohammad
 12/6/2019 2:40:15 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 06 05:43:07 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ120519CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 06 05:42:32 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.233	4.362	22747885	24633011	5.679	5.282
2)	SA Decachlor...	11.397	9.810	113.0E6	86724920	10.770	10.333
Target Compounds							
26)	L6 AR-1254-1	7.458	6.530	27128258	31056182	119.092m	110.994
27)	L6 AR-1254-2	7.688	6.687	42614567	28563525	116.401	113.991
28)	L6 AR-1254-3	8.070	7.116	44846850	48279706	112.466	108.468
29)	L6 AR-1254-4	8.364	7.355	37864246	30915208	116.735m	109.204
30)	L6 AR-1254-5	8.794	7.789	40754995	47195414	109.592m	109.427

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ120519\
Data File : PQ045657.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2019 01:27
Operator : SM\AJ
Sample : AR1254ICC100
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
AR1254101

Manual Integrations
APPROVED

mohammad
12/6/2019 2:40:15 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 06 05:43:07 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ120519CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 06 05:42:32 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

