

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ120524\
 Data File : PQ069597.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Dec 2024 15:09
 Operator : YP\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 05 22:18:03 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ120524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 05 22:16:15 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.418	2.790	361.9E6	495.5E6	75.528	73.072
2) SA Decachlor...	8.537	7.573	231.6E6	491.6E6	74.377	73.587
Target Compounds						
21) L5 AR-1248-1	4.502	3.795	69448378	105.0E6	731.130	716.116
22) L5 AR-1248-2	4.767	4.021	97629368	154.0E6	735.260	716.718
23) L5 AR-1248-3	4.955	4.058	115.5E6	150.2E6	735.351	718.395
24) L5 AR-1248-4	5.342	4.217	128.1E6	185.9E6	737.163	720.667
25) L5 AR-1248-5	5.377	4.593	127.5E6	187.7E6	755.870	722.324



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ120524\
Data File : PQ069597.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Dec 2024 15:09
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 05 22:18:03 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ120524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 05 22:16:15 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.418 min
Delta R.T.: 0.003 min
Response: 361915459
Conc: 75.53 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750




#1 Tetrachloro-m-xylene

R.T.: 2.790 min
Delta R.T.: 0.000 min
Response: 495525214
Conc: 73.07 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.537 min
Delta R.T.: 0.004 min
Response: 231552692
Conc: 74.38 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.573 min
Delta R.T.: 0.001 min
Response: 491609720
Conc: 73.59 ng/ml



#21 AR-1248-1


R.T.: 4.502 min
Delta R.T.: 0.002 min
Response: 69448378
Conc: 731.13 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750




#21 AR-1248-1

R.T.: 3.795 min
Delta R.T.: 0.000 min
Response: 104964841
Conc: 716.12 ng/ml




#22 AR-1248-2

R.T.: 4.767 min
Delta R.T.: 0.003 min
Response: 97629368
Conc: 735.26 ng/ml



#22 AR-1248-2


R.T.: 4.021 min
Delta R.T.: 0.001 min
Response: 153970943
Conc: 716.72 ng/ml



#23 AR-1248-3


R.T.: 4.955 min
Delta R.T.: 0.003 min
Response: 115544230
Conc: 735.35 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750




#23 AR-1248-3

R.T.: 4.058 min
Delta R.T.: 0.000 min
Response: 150195480
Conc: 718.39 ng/ml




#24 AR-1248-4

R.T.: 5.342 min
Delta R.T.: 0.002 min
Response: 128088646
Conc: 737.16 ng/ml



#24 AR-1248-4


R.T.: 4.217 min
Delta R.T.: 0.000 min
Response: 185889572
Conc: 720.67 ng/ml



#25 AR-1248-5

R.T.: 5.377 min
Delta R.T.: 0.003 min
Response: 127524135
Conc: 755.87 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750



#25 AR-1248-5

R.T.: 4.593 min
Delta R.T.: 0.000 min
Response: 187694945
Conc: 722.32 ng/ml